



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : etenders_ep@kmscl.kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

Equipment Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/323/2019

Dated: 06.08.2019

NOTICE

Sub:- Tender No.T323/2019- Results of technical document evaluation/ demonstration & Date of Price Bid Opening – Two Channel Clinical Audiometer - reg:-

- Ref:-**
1. Tender Inviting Notice published dated 01.04.2019
 2. Technical bid opening on 04.04.2019 & 07.04.2019@ 11.00 A M
 3. Clarification notice published on 15.05.2019
 4. Eligible list of demonstration was published on 27-05-2019
 5. Report of Demonstration/ document evaluation of Two Channel Clinical Audiometer Received on 30.07.2019

I. The reports of technical document evaluation of Two Channel Clinical Audiometer as per reference no. 5th cited above submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidders	Selected Model
1.	Two Channel Clinical Audiometer	M/s Alliance Biomedica Pvt.Ltd	GSI Audiostar Pro/Grason-Stadler (GSI), U.S.A
		M/s Auditivo Hearing Services Private Limited.	MA 42/Maico

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below:

A. KMSCL/EP/T323/1434/2019 : Two Channel Clinical Audiometer

1. M/s Alps International Pvt. Ltd

Rejected Model: AD629 (Extended Version) with FF & VRA Device/Interacoustics

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
4.	Provision of direct printout of the reports as PDF on SD-Card or USB.	This Feature is not available
	Comments	Not a Two channel Audiometer . It has only one interruptor switch.

2. M/s Pika Medical Private Limited

Rejected Model: Piano Plus VRA/Inventis, Italy

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
4.	Provision of direct printout of the reports as PDF on SD-Card or USB.	This Feature is not available
	Comments	Intensity unstable while changing frequencies. If intensity is stable Tone cannot be presented through both channels.

IV. The online price bid of the eligible tenderer mentioned above will be opened on 07 .08.2019 @ 11.00 am at the e-tender portal.

Dr. Dileep Kumar S R

കെ.എം.എസ് .സി.എൽ