



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [etenders\\_ep@kmscl.kerala.gov.in](mailto:etenders_ep@kmscl.kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

**Equipment Division**

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

**No: KMSCL/EP/321/2019**

**Dated: 20.08.2019**

**NOTICE**

**Sub:-** Tender No.T321/2019- Results of technical document evaluation/ demonstration & Date of Price Bid Opening – various equipments- reg:-

- Ref:-**
1. Tender Inviting Notice published dated 01.04.2019
  2. Technical bid opening on 03.04.2019 & 06.04.2019 @ 11.00 A M
  3. Clarification notice published on 08.05.2019 & 04.06.2019
  4. Eligible list for demonstration published on 27.05.2019 & 06.06.2019
  5. Demonstration/ document evaluation of various equipments conducted on 19.08.2019

I. The reports of technical document evaluation of various equipments as per reference no. 5th cited above submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidders	Selected Model
1.	Binocular Microscope Type III	M/s Universal Agencies	Vision 2000/ N.K. Jain Instruments
		M/s Scientific Medical Equipments (Kerala) Pvt	KXi2000/ Labex KK International
		M/s Lawrence And Mayo	DM-R L (DM-R Series)/ Debroy Engineers

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below:

**A. M/s Radical Scientific Equipments Pvt. Ltd**

Rejected Model: RXL-4/ Radical Scientific

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
3.	Objectives- Parfocal, antifungus coated 4x, 10x, 40x and 100x (oil immersion) with semiplan achromatic correction for FOV of 18 mm.	Not parfocal/ FOV more than 18 mm

**B. M/s Alan Medical & Laboratory Products**

Rejected Model: KLM-105/ Kwalitty Micro Scientific

Reason for rejection:



Clause No	Technical Specification	Deviations Noted
1	Inter changeable, inclined Binocular body, 360 degree rotatable head.	very loose head, falling down
2	Eyepieces - Highest quality 10 X wide field anti fungus field eyepiece. Comments	Eye Piece quality very poor
3	Objectives- Parfocal, antifungus coated 4x, 10x, 40x and 100x (oil immersion) with semiplan achromatic correction for FOV of 18 mm.	Not Parfocal no achromatic correction for FOV
4	Sub stage - Abbe condenser N.A 1.25 focusable, continuously variable iris diaphragm	Condenser not show N.A 1.25
6	Finish - A durable textured acid resistant finish	Quality Poor
	Comments	The fine adjustment is very close to coarse creating injury to finger while adjusting

**C. M/s Ast Meditech**

Rejected Model:Mi - Mi - xcl.b/ Micro Instrument company

Reason for rejection: Sample not produced

III. Due to administrative reason the tender for KMSCL/EP/T321/157/2019: ESR Analyzer is cancelled

IV. The online price bid of the eligible tenderer mentioned above will be opened on 22.08.2019 @ 11.00 am at the e-tender portal.

**Dr. Dileep Kumar S R**

കെ.എം.എസ് .സി.എൽ