

**KERALA MEDICAL SERVICES CORPORATION LTD**

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-3045647

0471-2337353, 0471-3045600

E-mail id : [etenders\\_ep@kmscl.kerala.gov.in](mailto:etenders_ep@kmscl.kerala.gov.in)

CIN : U24233KL2007SGC021616

**Equipment procurement Division****[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)****No: KMSCL/EP/526/2017****Dated: 08.06.2017****E-TENDER – PRE-TENDER MEETING – T230- AMENDMENT NOTICE****Sub:-** e-Tender no. 230/2017 released on 12/05/2017 – amendments published – reg.**Ref:-** Discussions in the pre-tender meeting held on 26/05/2017

The following amendments are made in respect of the tender clauses noted herein on the basis of the request from prospective tenderers, expert opinion and market study.

**TENDER NO. KMSCL / EP / T230 /136A/ 2017 : Ventilator CPAP ( Non invasive )**

Section No.	Clause No.	Existing clause	Amended clause
Appendix 1	1	Should have CPAP/ Auto CPAP mode with advanced sensor technology for tracking respiratory status of the patient	Should have Manual and Auto CPAP mode with advanced sensor technology for tracking respiratory status of the patient
Appendix 1	12	Added Clause	Should have ramp settings for CPAP Pressure.
Appendix 1	13	Added Clause	Should have mask size check feature
Appendix 1	14	Added Clause	Should have CPAP Pressure (4-20) cm of H <sub>2</sub> O
Appendix 1	15	Added Clause	Should have colour display of atleast 2 inches
Appendix 1	16	Added Clause	Should have inbuilt battery backup of 1 hour

**Dr. Dileep Kumar S R**  
General Manager