



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : etenders_ep@kmscl.kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

Equipment Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/98/2018

Dated: 09.08.2018

NOTICE

Sub:- Tender No. T266/2018- Results of technical document evaluation/ demonstration & Date of Price Bid Opening – Video Gastroscope & Colonoscope Model A - reg.

- Ref:-**
1. Tender Inviting Notice published dated 14.02.2018
 2. Technical bid opening on 26.03.2018@ 11.00 A M
 3. Clarification notice published on 03.05.2018
 4. Document evaluation of Video Gastroscope & Colonoscope Model A Received on 08.08.2018

I. The reports of technical document evaluation/ demonstration of Video Gastroscope & Colonoscope Model A as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderer qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Model / OEM
1.	Video Gastroscope & Colonoscope Model A	M/s. Cardinal medical technologies pvt ltd	EPK -3000(L) +EG-290KP +EG - 380LKP
		M/s. SOUTH INDIA SURGICAL CO. LTD	EG 250WR5 + EC 250WL+ EP x 2500+LMD 1530 MD

II. The items offered by the other tenderers participated in the tender are rejected due to the reasons mentioned against their name below.

1. M/s. Lakshmi Surgicals

Rejected Model : Karl Storz Video Gastroscope & Colonoscope Model A

Reason for rejection :

Clause No	Technical Specification	Deviations Noted
III,4	Should have in-built 150 W halogen light source or xenon.	No,100 W
III,5	Should have a stand-by lamp and should glow in the event of failure of the primary lamp.	No Stand-by lamp

III.The online price bid of the eligible tenderer mentioned above will be opened on 13.08.2018 @ 11.00 am at the etender portal.

Dr. Dileep Kumar S R
General Manager