

**KERALA MEDICAL SERVICES CORPORATION LTD***(A Government Of Kerala Undertaking)*

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : etenders_ep@kmscl.kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

Equipment Divisionwww.kmscl.kerala.gov.in

No: KMSCL/EP/321/2019

Dated: 07.09.2019

NOTICE

Sub:- Tender No.T321/2019- Results of technical document evaluation/ demonstration & Date of Price Bid Opening – various equipments- reg:-

- Ref:-**
1. Tender Inviting Notice published dated 01.04.2019
 2. Technical bid opening on 03.04.2019 & 06.04.2019 @ 11.00 A M
 3. Clarification notice published on 08.05.2019
 4. Eligible list for demonstration published on 27.05.2019 & 06.06.2019
 5. Demonstration/ document evaluation of various equipments conducted on 06.09.2019

I. The reports of technical document evaluation of various equipments as per reference no. 5th cited above submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidders	Selected Model
1.	Urine Analyzer	M/s Avantor Performance Materials India Limited	UA 220/Urit
		M/s Aspen Diagnostics Pvt Ltd	U-120 (Uriplus 1A)/Acon Biotec (Hangzhou)
		M/s Agappe Diagnostics Ltd	Mispa Uriskan 100 Plus/ Wuxi Opulen Madical Technologies Co.,Ltd
		M/s LabX	Uriscan Optima/YD Diagonostics CORP

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below:

1. KMSCL/EP/T321/1060/2019: Urine Analyzer**A. M/s Transasia Bio Medicals Ltd**

Rejected Model: Laura SMART/ Erba Lachema S.R.O

Reason for rejection: Sample not produced

B. M/s Siemens Healthcare Pvt Ltd

Rejected Model: Siemens Clinitek Status Plus/ Siemens Healthcare Diagnostics Inc

Reason for rejection:

Clause No	Technical Specification	Deviations noted
1.e	Glucose	No, On demonstration with samples the offered model showed a lower positivity for glucose.

C. M/s Wheecon Instruments Private Limited

Rejected Model: Uryxcon Relax/ Machery-Nagel GnbH&Co. Germany

Reason for rejection: Sample not produced

III. Due to administrative reason the tender for KMSCL/EP/T321/1060/2019: Hematology Analyzer (5 part) Model A is cancelled

IV. The online price bid of the eligible tenderer mentioned above will be opened on 16.09.2019 @ 11.00 am at the e-tender portal.

Dr. Dileep Kumar S R

കെ.എം.എസ് .സി.എൽ