



Running Contract Details	
Equipment Name	12 Channel Stress Test Equipment with Tread Mill
Running Contract Valid Till	28-11-2020
Tender Ref No	KMSCL/EP/T284/272/2018(R)
Tendered Quantity	35
Supplier Name	M/s Schiller Healthcare India Pvt. Ltd
GST No	34AADCS5091B1ZN
Model & Make	Cardiovit AT 170 / Schiller
Unit rate (Rs)	7,63,500.00
CGST 0.00%	0.00
SGST 0.00%	0.00
IGST 12.00%	91,620.00
Total Cost(Exl.KMSCL S.C) Rs.	8,55,120.00
Flood Cess (1%) Rs.	7,635.00
Service Charges 7% + GST 18%	63,065.10 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	9,25,820.10
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Schiller Healthcare India Pvt Ltd 37/1013 2nd Floor Unidec House Fathima Church Road Elamkulam Kochi-682020	Contact Person	Mr.Stijo Mathew , Mr. P.S Vidhyashankar
	Phone	04842203264
	Mobile No	09846706787, 09383620520
	Email	stijo@schillerindia.com, vidyashankar@schillerindia.com, mail@schillerindia.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	12 Channel Stress Test Equipment with Tread Mill	862755	63065.1	925820.1
		862755	63065.1	925820.1

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
12 Channel Stress Test Equipment with Tread Mill							
Labour	6,500.00	6,500.00	6,500.00	6,500.00	6,500.00	6,500.00	6,500.00
Comprehensive	32,500.00	32,500.00	32,500.00	32,500.00	32,500.00	32,500.00	32,500.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :12 Channel Stress Test Equipment with Tread Mill

1. Treadmill System – Should be computerized and noise less and not dependent on any other company for spares support
2. Display Technology – should be advanced & user –friendly – LED backlight
3. No. of leads display – 12 lead – simultaneous
4. System – Dedicated Microcomputer with Dedicated incremental updating
5. ECG Scanning Technique – Incremental updating Template matching or equivalent technology
6. Auto / Manual ST analysis
7. Post ‘j’ adjustments – all three E,J, & Post ‘j’ and point adjustments are possible with auto / manual control of base line
8. Arrhythmia documentation in stress mode –Y
9. ST segment monitoring – 12 leads or more
10. No. of leads displayed – 12 leads or more(real time) and should be displayed on the screen simultaneously
11. Signal in Acquisition Module – signal must be immune to the environmental noise.
12. Filtering Techniques – for advances correction of artifacts
13. Printer - Real time Thermal printer.

14. Protocols – Standard & minimum 10 numbers of user defined protocols.
15. Storage Media – data storage may be possible on HD, as well as CD archive.
16. Windows based OS
17. Facility for advanced networking
18. Should support standard protocols like Bruce, Modified Bruce, etc.
19. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.