



Running Contract Details	
Equipment Name	Advanced GI Endoscopy Electrosurgical instruments
Running Contract Valid Till	14-09-2019
Tender Ref No	KMSCL/EP/T231/974/2017
Tendered Quantity	4
Supplier Name	M/s Neoptic Surgicals
GST No	32ACFPK5938G1ZO
Model & Make	ERBE VIO/Erbe Electromedizin GmbH, Germany
Unit rate (Rs)	10,69,642.00
CGST 6.00%	64,178.52
SGST 6.00%	64,178.52
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	11,97,999.04
Flood Cess (1%) Rs.	10,696.42
Service Charges 7% + GST 18%	88,352.43 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	12,97,047.89
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
35/1476 Kripa South Janatha Road Palarivattom Kochi-682025	Contact Person	Mr. K.J Kurian
	Phone	
	Mobile No	9447039805
	Email	neopticsurgicals@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Advanced GI Endoscopy Electrosurgical instruments</b>	1208695.46	88352.43	1297047.89
		<b>1208695.46</b>	<b>88352.43</b>	<b>1297047.89</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Advanced GI Endoscopy Electrosurgical instruments</b>							
Labour	36,877.00	36,877.00	61,463.00	61,463.00	61,463.00	73,755.00	73,755.00
Comprehensive	49,171.00	61,463.00	73,755.00	86,048.00	98,340.00	1,10,633.00	1,22,925.00

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Advanced GI Endoscopy Electrosurgical instruments**

1. The unit should be a Microprocessor controlled 200 watts class Electro surgery unit with TFT display
2. The unit should have 1 Bipolar and 1 Monopolar sockets with universal sockets.
3. The unit should also have facility to modify or add additional sockets in field based on requirements of the doctor.
4. The unit should have internal cooling system (no conventional external air cooling system)
5. The unit should have two different fractionated cut modes for Endoscopy work.
6. The unit should have facility to modulate the fractionated cut mode's cut interval and cut duration.
7. The unit should have bipolar cut mode and bipolar coagulation modes. (Auto start and Auto stop functions)
8. To unit should be upgradable to APC and water jet to do advance procedures in future
9. The unit should be a updatable and upgradable unit
10. Type : protection class CF, class 1, IIb and CE and FDA approved
11. The unit should have continuous neutral electrode monitoring system and Neo- Natal safety system for safer applications even with neonatal patients
12. The unit should have programmable facility for customizing 198 programs
13. The unit should also be supplied with standard accessories such as:
  - a. Two pedal foot switch
  - b. Universal Monopolar cable
  - c. Patient plate disposable – 50 no's
14. All other accessories required for making the unit completely functional shall be supplied.