



Running Contract Details	
Equipment Name	Anaesthesia Workstation Model B
Running Contract Valid Till	13-12-2020
Tender Ref No	KMSCL/EP/T288/2B/2018(R)
Tendered Quantity	100
Supplier Name	M/s Sree Gokulam Healthcare Pvt Ltd
GST No	32AAJCS3502K1ZI
Model & Make	Wato EX-55 Pro / Shenzhen Mindray Bio- Medical Electronics Co.Ltd
Unit rate (Rs)	20,35,000.00
CGST 6.00%	1,22,100.00
SGST 6.00%	1,22,100.00
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	22,79,200.00
Flood Cess (1%) Rs.	20,350.00
Service Charges 7% + GST 18%	1,68,091.00 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	24,67,641.00
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Dhanya Complex Sastha Temple Road Ernakulam North PO Cochin-682018	Contact Person	Anoop
	Phone	
	Mobile No	9995813333
	Email	sales@gokulamhealthcare.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Anaesthesia Workstation Model B	2299550	168091	2467641
2	Vaporizer - Isoflurane	87575	6401.5	93976.5
3	Vaporizer - Sevoflurane	87575	6401.5	93976.5
4	Continious Cardiac output measurement facility	146900	10738	157638

Item-wise Price Details			
	by PICCO / LIDCO technology with all accessories		
		2621600	191632
			2813232

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Anaesthesia Workstation Model B							
Labour	85,000.00	85,000.00	85,000.00	85,000.00	85,000.00	85,000.00	85,000.00
Comprehensive	2,15,000.00	2,25,750.00	2,37,038.00	2,48,889.00	2,61,300.00	2,74,000.00	2,88,120.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment : Anaesthesia Workstation Model B

I. Operational Requirements

1. Anaesthesia machine complete and integrate with anesthesia gas delivery system; Circle absorber system; Precision vaporizer for isoflurane and sevoflurane; Anesthesia Ventilator, Monitoring system To monitor anesthesia gases, ECG, EtCO₂, FiO₂ (online O₂ Analyzer), Pulse Oximeter and airway pressure, NIBP, IBP,

1. Rectal & skin temperature.
2. Essential accessories to make the system compete and compatible with the existing system of gas outlet.

II. Flow management

1. Should be compact, ergonomics & easy to use.
2. Machine should provide with electronic gas mixing
3. Integrated Multi-Color TFT display of at least 15" size, with virtual flow meter for O₂, N₂O – or Air
4. Single/ dual flow sensing capability at exhalation and inhalation port.

5. Should have back up O₂ control which provides' an independent fresh gas source and flow meter control in case of electronic failure (Auxillary flowmeter).
6. One number yoke each for O₂ and N₂O. Separate pipeline inlet for oxygen, Nitrous Oxide and Air.
7. Hypoxic Guard to ensure minimum 25% O₂ across all O₂-N₂O mixtures and oxygen failure warning.

III. Breathing System

1. Latex free fully auto cleavable
2. Flow sensing capability at inhalation or exhalation ports, sensor connections shall be internal to help prevent disconnect
3. Sensor should not require daily Maintenance
4. Bag to vent switch shall be bi-stable and automatically begins mechanical ventilation in the ventilator position
5. Adjustable pressure limiting valve shall be flow and pressure compensated

IV. Standard circle absorber system

1. Should have adjustable pressure limiting valve, breathing circuit pressure measuring device
2. Should have a bag/ventilator selecting valve integrated onto the absorber
3. Should be suitable to use low flow techniques
4. Facility to attach oxygen sensor
5. Should have CO₂ absorbent chamber canister with CO₂ bypass

V. Vaporizers

1. New generation vaporizer must be isolated from the gas flow in the off position and prevent the simultaneous activation of more than one vaporizer.
2. Vaporizer should mount to a selectatec or equivalent manifold of to vaporizers, which allows easy exchange between agents. Temperature, pressure and flow compensated vaporizers and maintenance free — for isoflurane and Sevoflurane

VI. Ventilator (integrated)

1. The workstation should have integrated anesthesia ventilator system for adult and pediatric.
2. Ventilator should have volume control and pressure controlled SIM V/P, PRVC/ PCVVG/ Auto flow and PEEP.
3. Ventilator should have a tidal volume compensation capability to adjust for losses due to compression, compliance and leaks; and compensation for fresh gas flow.

4. The workstation should be capable of delivery of low flow anesthesia.

5. Ventilator should be capable of at least 110L/min peak flow.

VII. Anesthesia monitoring system should be modular:

1. Monitoring of vital parameters; ECG (5 Leads) with ST segment analysis, NIBP, SPO2 and 2 invasive blood pressure & Spirometry with display of flow volume loop (Either in Ventilator or Patient Monitor). Monitor size should be atleast 15" touch screen..

2. Twin temperature measurement with skin and rectal probes - Two set with each monitor.

3. Automatic identification and measurement of anesthetic agents, EtCO2, O2, FiCO2, N2O, MAC value FiO2 and FeO2 measurement (Should work either in Ventilator or in patient monitor).

4. Depth of Anesthesia monitoring module BIS/ ENTROPY - one per monitor with 20 sensors with minimum 10 months shelf life.

5. Neuromuscular monitoring with all accessories.

6. Continuous Cardiac output measurement facility by PICCO / LIDCO technology with all accessories. The rates shall be quoted separately in the BOQ.

7. 24 Hours of graphical and numerical trending.

8. Should have hemodynamic, Oxygenation and ventilation. Calculation package.

9. Should have a detachable monitor module that serves as a transit monitor/ separate monitor with the following parameters SPO2, ECG and IPB.

10. Facility for storing snaps shot/ event recording during critical events for waveform review at a later stage.

11. Audio visual and graded alarming system.

VIII. Display of ventilator

1. Tidal volume (VT)

2. Inspiratory /Expiratory ratio (I:E).

3. Inspiratory pressure (P inspired)

4. Pressure limit (P limit)

5. Positive End Expiratory Pressure (PEEP)

6. Ventilator waveform

IX. The equipment should have the provision for Centralized monitoring and Networking

X. System Configuration Accessories, spares and consumables

1. Anaesthesia Gas Delivery system -01

2. Circle absorber -.01
3. Ventilator -01
4. Monitor -01
5. Vaporizer Sevoflurane -01
6. Vaporizer Isoflurane -01
7. Adult and Paediatric autoclavable silicone breathing circuit -02
8. Reuseable IBP cable – 2 nos and disposable IBP Transducer -10
9. Temp probe Skin reusable -02
10. Temp probe Rectal Reusable -02
11. Accessories Anaesthetic gases -01 set
 - a. sample line – 10 nos
 - b. water trap – 10 nos
12. Depth of Anaesthesia Sensors - 20
13. Accessories for Cardiac Output module -01 set
14. Accessories for neuromuscular transmission monitor -01 set
15. ECG 5 lead – 1 No, SPO2 Reuseable Adult – 1 No, NiBP tubes and cuffs 3 sizes (Medium, large and Extra large)
16. Disposable Adult & Pediatric circuits -50 each
17. HME filters -25 Nos
18. Microstream / Side stream ETCO2 disposable kit for adult-25 nos, paediatric – 2 nos.
19. Should have retractable/ foldable writing tray.

XI. Environmental factors

1. Environmental factors Machine should have facility to connect to active AGSS (Anaesthetic Gas Scavenging System/port) at the user institution if a working scavenging system provided by the user is available. The key plug for AGSS should be provided by the user institution. Should also supply passive scavenging tube.

XII. Power Supply

1. Power input to be 220-240VAC, 50HZ/ as appropriate fitted with Indian plug.
2. UPS of suitable rating shall be supplied / In built battery backup for minimum 1 hour for the entire system. Atleast two auxiliary power outlets should be available.

XIII. Standards, Safety and Training

1. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.

2. The Anaesthesia machine and Ventilator should be from one manufacturer
3. Certificate of calibration and inspection from factory shall be provided
4. Should supply with 5 kg Soda Lime along with machine