



Running Contract Details	
Equipment Name	Dental Lathe
Running Contract Valid Till	02-05-2020
Tender Ref No	KMSCL/EP/T257/404F/2017
Tendered Quantity	100
Supplier Name	M/s Confident Dental Equipments Pvt Ltd
GST No	32AAACC9198C1ZS
Model & Make	Labo21 / Confident
Unit rate (Rs)	11,500.00
CGST 6.00%	690.00
SGST 6.00%	690.00
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	12,880.00
Flood Cess (1%) Rs.	115.00
Service Charges 7% + GST 18%	949.90 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	13,944.90
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
No.4 - 1st Floor Shivdas Tower Opp. Maymoon Theatre Chittoor Road Cochin - 11	Contact Person	Mr.sarath,Mini
	Phone	
	Mobile No	9995592737,8089383364
	Email	cdelcochin@vsnl.net,marketingmngr@confidental.org,surgident@vsnl.net

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Dental Lathe	12995	949.9	13944.9
		12995	949.9	13944.9

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Dental Lathe

1. Should have continuous rating having adjustable speed minimum of 1500 RPM.
2. Should be a table top model
3. Should have vibration free operation with low noise level
4. It should have the arrangement for buffing cone/ drill check /Grinding Stones, 3chuck buffer, stone & bur chucks.
5. Should operate on mains 220-240Vac, 50 Hz single phase