



Running Contract Details	
Equipment Name	Electronic Weighing Scale for Blood Bank
Running Contract Valid Till	06-12-2019
Tender Ref No	KMSCL/EP/T239/267A/2017
Tendered Quantity	20
Supplier Name	M/s Scientific Enterprises
GST No	32AAIFS7780D1Z4
Model & Make	HT-BBW1 / Tulabhavan
Unit rate (Rs)	19,475.00
CGST 9.00%	1,752.75
SGST 9.00%	1,752.75
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	22,980.50
Flood Cess (1%) Rs.	194.75
Service Charges 7% + GST 18%	1,608.64 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	24,783.89
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
PB No. 1951 Vytila Kochi â€“ 682019	Contact Person	Deepa
	Phone	
	Mobile No	9447717286
	Email	Seqcochin@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Electronic Weighing Scale for Blood Bank	23175.25	1608.64	24783.89
		23175.25	1608.64	24783.89

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 rd Year	4 th Year	5 th Year
Electronic Weighing Scale for Blood Bank			
Labour	1,500.00	1,500.00	1,500.00
Comprehensive	2,500.00	2,500.00	2,500.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Electronic Weighing Scale for Blood Bank

1. Resolution: 1gm/1ml
2. Transducer: Load cell
3. Keyboard : Weight/Volume conversion Volume setting (Up/dow)
4. Ingress of water: IPX1, protected against dripping water
5. Weighing range: Up to 2.5kg
6. Interface facility with plasma: Standard TTL interface with Plasma expressor
7. Zero set provision: Automatic on initialization/manual
8. Input supply:230+/-10% VAC, 50Hz
9. Switch: Power ON/OFF
10. Indications: zero set
:gm/ml
: Over range
: Indication when set
:Weight/Volume is achieved

11. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.

12. Manufacturing standard:ISO 9001:2008