



Running Contract Details	
Equipment Name	Flexible Naso-pharygo Laryngoscope
Running Contract Valid Till	27-12-2020
Tender Ref No	KMSCL/EP/T288/46/2018(R)
Tendered Quantity	150
Supplier Name	M/s Lakshmi Surgicals
GST No	32AACFL2603A1ZB
Model & Make	Karl Storz / Karl Storz GmbH & Co
Unit rate (Rs)	4,51,217.95
CGST 6.00%	27,073.08
SGST 6.00%	27,073.08
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	5,05,364.10
Flood Cess (1%) Rs.	4,512.18
Service Charges 7% + GST 18%	37,270.60 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	5,47,146.88
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
203XIX -City Towers Bank Jn -Market Road Aluva -Ernakulam - 683101 Kerala	Contact Person	Mr. Praveen
	Phone	
	Mobile No	9995271784
	Email	lakshmisurgicals2005@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Flexible Naso-pharygo Laryngoscope	509876.28	37270.6	547146.88
2	Light Source	134554.42	9581.2	144135.62
3	Battery operated portable Light Source	31361.11	2292.41	33653.53
4	Camera System	482402.13	30888.69	513290.83

Item-wise Price Details					
5	LED TV		24990	1734.6	26724.6
6	Trolley		19867.05	1379.01	21246.06
			1203050.99	83146.52	1286197.52

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Flexible Naso-pharyngo Laryngoscope							
Labour	7,549.62	7,549.62	7,549.62	7,549.62	7,549.62	7,549.62	7,549.62
Comprehensive	20,132.33	20,132.33	20,132.33	20,132.33	20,132.33	20,132.33	20,132.33

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Flexible Naso-pharyngo Laryngoscope

1. Should have a field of view of at least 75 degree.
2. Should have a depth of field from 5 to 50 mm.
3. The insertion tube should have maximum (3- 3.8) mm diameter or less.
4. Should have at least 130 degree upwards and 100 degree downwards angulations.
5. Should have a working length of at least 300 mm.
6. Should have a light guide illuminating system.
7. Should provide suitable light source / battery operated portable type.
8. Should be supplied with all standard accessories including storage box , pressure compensation cap, cleaning brush, Biopsy forceps, Grasping forceps and leakage tester.
9. Should have a working channel diameter of at least 1.2mm – 1.5mm

II. Light Source:

1. Should have an LED light source equivalent to 150W light output.
2. Should have manual light intensity control.
3. Should have a cooling system.
4. Should work with input 200 to 240 Vac 50Hz supply.
5. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid

III. Camera system

1. Should be a single chip camera technology.
2. Should have one composite video outputs and one S-video output.
3. Should have anti-moisture filter for fiber scopes.
4. Should have fully automatic exposure control.
5. Should have automatic white balance with memory function.
6. Should have horizontal resolution of more than 450 lines
7. Should be supplied with suitable coupler and video cables
8. Should work with input 200 to 240Vac 50 Hz supply.
- 9.

IV. LED TV of 21 inches or higher

V. Trolley

1. Should be supplied with suitable trolley.
2. Trolley should have at least 5 power sockets to connect the camera, light source, monitor etc.