



Running Contract Details	
Equipment Name	Infant CPR Manikin
Running Contract Valid Till	07-03-2021
Tender Ref No	KMSCL/EP/T307/844/2019(R)
Tendered Quantity	276
Supplier Name	M/s Laerdal Medical India Pvt Ltd
GST No	33AABCL8105F1ZX
Model & Make	Resusci Baby / Laerdal Medical
Unit rate (Rs)	80,000.00
CGST 9.00%	7,200.00
SGST 9.00%	7,200.00
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	94,400.00
Flood Cess (1%) Rs.	800.00
Service Charges 7% + GST 18%	6,608.00 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	1,01,808.00
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
No.10 1st Floor 1st Street Dr. Subbarayan Nagar Kodambakam Chennai- 600 024.	Contact Person	Balaganesh
	Phone	044-42614773
	Mobile No	9840153367
	Email	balaganesh.mahendrathas@laederal.co.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Infant CPR Manikin	95200	6608	101808
		95200	6608	101808

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Infant CPR Manikin

1. The manikin should be a low-cost, life like infant CPR trainer
2. Should have realistic chest compressions and rise that allow students to learn proper techniques
3. Must have oral and nasal passages that allow realistic nose pinch required for mouth-to-nose ventilation
4. Must have natural obstruction of the airway that allows students to learn the important technique of opening the airway
5. Must have head tilt/chin lift and jaw thrust that allows students to practice correctly all manoeuvres necessary when resuscitating a real patient
6. Should have a realistic chest compliance that means students can experience the proper technique required for chest compression on infants
7. Must have foreign – body airway obstruction feature that allows the release of a foreign-body obstruction to be practiced through back blows and chest-thrust techniques
8. Must be supplied with economical disposable airways for quick and easy clean up
9. Should be supplied with removable and reusable faces for convenient and affordable maintenance.
10. The individual manikin configuration should come standard with a carry Bag which also serves as a padded training mat.
11. The manikin must be equipped with means of providing feedback on performed CPR, compression recoil depth, rate, rhythm, interpretation time& ventilation volume on a battery operated audio visual electronic feedback unit. Models supplied with feedback unit/manikin that requires 230V power connection will not be accepted.