



Running Contract Details	
Equipment Name	Infusion Pump
Running Contract Valid Till	07-11-2020
Tender Ref No	KMSCL/EP/T275/62/2018
Tendered Quantity	625
Supplier Name	M/s Anamdev Engineers
GST No	32ABKPG3145P1ZJ
Model & Make	Graseby1200 / Smiths Medical
Unit rate (Rs)	22,950.00
CGST 6.00%	1,377.00
SGST 6.00%	1,377.00
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	25,704.00
Flood Cess (1%) Rs.	229.50
Service Charges 7% + GST 18%	1,895.67 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	27,829.17
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
1st Floor V̄rindavan Door No. 43/1606 B North Railway Station Road Ernakulam Cochin-682018	Contact Person	Mr. Jayakumar
	Phone	
	Mobile No	9847049686
	Email	mail@anamdev.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Infusion Pump	25933.5	1895.67	27829.17
		25933.5	1895.67	27829.17

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 rd Year	4 th Year	5 th Year
Infusion Pump			
Labour	1,200.00	1,275.00	1,350.00
Comprehensive	1,950.00	2,100.00	2,250.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Infusion Pump

1. Should be operated on drip rate Peristaltic pump method.
2. Should compatible with most of the IV set (macro/micro drip sets). Should be supplied with 100 Nos of compatible IV sets.
3. Should have the following flow rates.
4. IV set ml/hr drops/min
15 drops/ml 3-450 ml/hr 1-100 drops/min
20 drops/ml 3-450 ml/hr 1-100 drops/min
60 drops/ml 1-100 ml/hr 1-100 drops/min
5. Should have a flow rate accuracy of $\pm 10\%$ and drip rate accuracy of $\pm 2\%$.
6. Should have a volume infused display from 0 to 999.9ml.
7. Should have a purge and KVO facility.
8. Should have an audible and visual alarm for occlusion pressure, air alarm, door open, empty, low battery, free flow and disconnection.
9. Should have a LCD graphical display with backlight and graphical display of infusion. Should have a minimum 4hr battery back up at highest delivery rate.
10. Should work with input 200 to 240ac 50Hz supply.

11. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid

12. Should have different modes of infusion, time, rate and dose.