



Running Contract Details	
Equipment Name	Microdebrider with sinus shaver blades and mastoid drills
Running Contract Valid Till	10-09-2019
Tender Ref No	KMSCL/EP/T226/999/2016
Tendered Quantity	6
Supplier Name	M/s Neoptic Surgicals
GST No	32ACFPK5938G1ZO
Model & Make	Stryker USA
Unit rate (Rs)	3,44,355.36
CGST 6.00%	20,661.32
SGST 6.00%	20,661.32
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	3,85,678.00
Flood Cess (1%) Rs.	3,443.55
Service Charges 7% + GST 18%	28,443.75 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	4,17,565.31
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
35/1476 Kripa South Janatha Road Palarivattom Kochi-682025	Contact Person	Mr. K.J Kurian
	Phone	
	Mobile No	9447039805
	Email	neopticsurgicals@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Microdebrider with sinus shaver blades and mastoid drills	389121.55	28443.75	417565.31
2	Debrider Hand Piece	268850.21	19652.24	288502.44
3	Foot Switch	62040.03	4534.96	66574.99

Item-wise Price Details				
4	4mm,11cm, 40 degree aggressive balde	9534.38	696.94	10231.31
5	4mm, 8mm aggressive serrated straight blade	3707.81	271.03	3978.84
6	60degree or 70 degree cutting blade	7945.31	580.78	8526.09
7	4mm, 11cm, 12 or 15 degree cutting blade	7945.31	580.78	8526.09
8	Hooded burr 3mm, 8cm	8475	619.5	9094.5
9	Adenoidectomy Blade	7945.31	580.78	8526.09
		765564.91	55960.76	821525.67

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Microdebrider with sinus shaver blades and mastoid drills							
Labour	30,378.00	30,378.00	36,454.00	36,454.00	39,491.00	42,530.00	42,530.00
Comprehensive	33,753.00	33,753.00	40,504.00	40,504.00	43,878.00	47,254.00	47,254.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Microdebrider with sinus shaver blades and mastoid drills

I. Console

1. Electrical powered system should be compatible to operate different kinds of hand pieces like debrider H/P and mastoid drill. Console should recognize different hand pieces automatically.
2. The system should be user friendly and the parameters should be able to adjust from the control panel and the display should suggest operating parameters for a variety of ENT procedures.
3. Should have a inbuilt irrigation pump.
4. The various functions like speed/mode- forward or reverse of active hand piece should be able to adjust control panel.
5. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid

I. Micro Debrider Hand Piece

1. The debrider hand piece should be able to operate in forward at speeds upto 12,000 rpm and 5,000 rpm in oscillating mode.
2. Should have selection for rotation (clockwise / anticlockwise/ oscillation).
3. The debrider hand piece should be capable of accepting various types of blades and burrs.
4. Should be able to adjust the irrigation level of burr/ blade with the touch of button.
5. The debrider should have blades for all adenoidectomy and tonsillectomy procedures.
6. The debrider should support burrs required for frontal sinus , DCR procedures.
7. Sterilizable through ETO / Steam and flash autoclavable.
8. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid.

I. Tools And Blades

1. 4 mm double serrated straight blade - 5 Nos.
2. 4mm, 40 degree double blade - 3 Nos.
3. 60 to 70 degree cutting blade - 2 Nos.
4. 4mm , 11 cm, 12 to 15 degree cutting blade - 2 Nos.
5. Adenoidectomy Blade - 3 Nos.
6. Hooded burr (3 – 4) mm, (8- 12) cm - 2 Nos.

I. Items to be Quoted As Standard Accessories

1. Electrical console with pump and accessories - 1 No.
2. Debrider hand piece - 1 No.
3. Foot switch - 1 No.