



Running Contract Details	
Equipment Name	Micromotor for ENT Model B
Running Contract Valid Till	21-06-2020
Tender Ref No	KMSCL/EP/T266/1174/2018
Tendered Quantity	60
Supplier Name	M/s Stryker India Private Ltd.
GST No	06AAECS2513F1ZS
Model & Make	micromotor system (saber drill) / stryker instrument
Unit rate (Rs)	7,09,505.00
CGST 0.00%	0.00
SGST 0.00%	0.00
IGST 12.00%	85,140.60
Total Cost(Exl.KMSCL S.C) Rs.	7,94,645.60
Flood Cess (1%) Rs.	7,095.05
Service Charges 7% + GST 18%	58,605.11 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	8,60,345.76
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
10th Floor Vatika Business Park Block II Sohna Road Sector 49 Gurgaon-122002 Haryana	Contact Person	Savita Saini
	Phone	01244850500,01244850600
	Mobile No	9910402662
	Email	tenders.india@stryker.com,savita.saini@stryker.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Micromotor for ENT Model B	801740.65	58605.11	860345.76
		801740.65	58605.11	860345.76

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Micromotor for ENT Model B							
Labour	8,001.00	8,802.00	9,682.00	10,650.00	11,715.00	12,886.00	14,175.00
Comprehensive	21,337.00	21,337.00	26,672.00	26,672.00	32,006.00	32,006.00	37,340.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Micromotor for ENT Model B

1. Should have a micro motor with variable speed from 10,000 to 60,000rpm.
2. Should have at least four speed control modes.
3. Should have an brushless dc motor for continuous high speed performance
4. Should have ON and OFF control for console and speed acceleration foot control
5. Should have integrated irrigation pump
6. Should be supplied with straight hand piece for micro motor and contra angle hand piece for micro motor.
7. Should be supplied with the following Burs set.
 - Tungsten Carbide burs - 10 nos. (Assorted sizes)
 - Diamond burs - 10 nos. (Assorted sizes)
 - Steel cutting burs - 10 nos. (Assorted sizes)
9. Should work with input 200 to 240Vac 50 Hz supply.
10. Should have forward and reverse mode.
11. Should have safety certificate from a competent authority CE/FDA (US). Other equivalent certificates will not be accepted. Should have valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid

