



Running Contract Details	
Equipment Name	OT Table Electric - B
Running Contract Valid Till	19-01-2020
Tender Ref No	KMSCL/EP/T248/197B/2017(R)
Tendered Quantity	25
Supplier Name	M/s Trichur Surgicals
GST No	32AAFFT9094G1ZX
Model & Make	Warrior / Staan
Unit rate (Rs)	2,34,000.00
CGST 9.00%	21,060.00
SGST 9.00%	21,060.00
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	2,76,120.00
Flood Cess (1%) Rs.	2,340.00
Service Charges 7% + GST 18%	19,328.40 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	2,97,788.40
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Veliyannur Road Thrissur â€“ 21	Contact Person	CD David
	Phone	
	Mobile No	9447020813
	Email	triso@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	OT Table Electric - B	278460	19328.4	297788.4
		278460	19328.4	297788.4

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
OT Table Electric - B							
Labour	5,000.00	5,000.00	5,000.00	7,500.00	7,500.00	10,000.00	10,000.00
Comprehensive	2,000.00	2,000.00	2,000.00	3,000.00	3,000.00	3,000.00	4,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :OT Table Electric - B

1. The table should have minimum of 4 sections ie. head section, leg section, seat section and back plate section.
2. The table should be electrically operated having the following hand switch operated functions (all the dimensions will have a permitted deviation of +/- 10 %)

Description Range

- i. Up 1000mm Maximum
- ii. Down 680mm Minimum
- iii. Trendelenburg 25 deg
- iv. Reverse Trendelenburg 25 deg
- v. Right Lateral tilt 20 deg
- vi. Left Lateral tilt 20 deg
- Vii. Back up 80deg
- viii. Back down 40deg

3. In addition to the above motorized operated functions, the table must have the following manual functions.

a. Head section tilting

ii. The leg section should have 90° down movement and should move side wards to a minimum of 90 degree.

4. The table should have in-built kidney bridge

5. Should have enhanced weight bearing casters fitted with ball bearing.

6. Table should have a stable braking position with single lever foot operated brake pedal.

7. The table should be supplied with the following accessories.

i. Mattress for the complete table top in sections - 1 set

ii. A pair of arm boards with pad and fixing clamp – 1

iii. A pair of padded shoulder support with clamps (SS grade 304) – 1

iv. A pair of padded lateral support with clamps (SS grade 304) – 1

v. A pair of padded leg crutches with clamps (SS grade 304) – 1

vi. Anesthetic screen frame with clamp (SS grade 304) – 1

vii. Should supply Patient restraint strap – 1No, A pair of padded leg support to use in trendelenburg position.

viii. The table should have a heavy and sturdy base and compact to provide adequate foot room for the operating team.

ix. The base cover, lifting column cover and side rails should be made of stainless steel.

8. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid

9. The table shall have a radiolucent tabletop

10. Battery backup shall be available for minimum 50 functions.

11. Should have a minimum weight bearing capacity of 200Kg.