



Running Contract Details	
Equipment Name	Peripheral Nerve Stimulator
Running Contract Valid Till	30-01-2020
Tender Ref No	KMSCL/EP/T248/98/2017(R)
Tendered Quantity	52
Supplier Name	M/s Sanmed Healthcare
GST No	32AHFPV1256R1ZT
Model & Make	NS -100 / Inmed Equipments Pvt Ltd
Unit rate (Rs)	50,000.00
CGST 6.00%	3,000.00
SGST 6.00%	3,000.00
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	56,000.00
Flood Cess (1%) Rs.	500.00
Service Charges 7% + GST 18%	4,130.00 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	60,630.00
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
TC 66/468 - Pachaloor Thiruvallom Trivandrum - 695027	Contact Person	P.Vijay
	Phone	0471-2314248
	Mobile No	9995711652,8138911652
	Email	sales@sanmedgroup.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Peripheral Nerve Stimulator</b>	56500	4130	60630
		<b>56500</b>	<b>4130</b>	<b>60630</b>

<b>Annual / Comprehensive Maintenance Charges (Exl.Tax)</b>			
<b>Rate</b>	<b>3<sup>rd</sup> Year</b>	<b>4<sup>th</sup> Year</b>	<b>5<sup>th</sup> Year</b>
<b>Peripheral Nerve Stimulator</b>			
<b>Labour</b>	3,304.00	3,894.00	4,484.00
<b>Comprehensive</b>	8,260.00	8,850.00	8,850.00

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Peripheral Nerve Stimulator**

1. Should have a current range from 0 .1 to 99mA.
2. Should have a resolution of 1mA
3. Should have a digital display for current.
4. Should have short stimulus pulse duration of 0.1ms.
5. Should have a frequency range from 0.1 to 99 Hz.
6. Should be able to deliver single twitch, Train of four, PTC and DBC..
7. Should have integrated electrode cable with lead.
8. Should be battery operated with rechargeable battery.
9. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.