



Running Contract Details	
Equipment Name	Platelet Agitator Incubator
Running Contract Valid Till	02-06-2021
Tender Ref No	KMSCL/EP/T321/264/2019
Tendered Quantity	100
Supplier Name	M/s Kochin Inlab Equipments India Pvt Ltd
GST No	32AADCK3263F1ZX
Model & Make	PI 200 & PA 300 / Terumo Penpol Pvt Ltd
Unit rate (Rs)	1,70,000.00
CGST 6.00%	10,200.00
SGST 6.00%	10,200.00
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	1,90,400.00
Flood Cess (1%) Rs.	1,700.00
Service Charges 7% + GST 18%	14,042.00 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	2,06,142.00
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Door #28/2391-D Vihaan - Chilavannur Kadavanthra P.O Kochi-682020.	Contact Person	Manoj
	Phone	
	Mobile No	9447790938
	Email	manoj.inlab@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Platelet Agitator Incubator	192100	14042	206142
		192100	14042	206142

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Platelet Agitator Incubator							
Labour	3,000.00	3,000.00	3,000.00	3,000.00	3,000.00	3,000.00	3,000.00
Comprehensive	4,250.00	4,250.00	4,250.00	4,250.00	4,250.00	4,250.00	4,250.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Platelet Agitator Incubator

1. Capacity : 48 Platelets Bags holding Capacity
2. Temperature ; 220 C
3. Oscillation : 60-75 cycles per minute
4. Temperature controller ; Micro controller based Temperature controller
5. Agitator : 230V AC induction motor with agitating mechanism
6. Outer body : MS powder coated
7. Inner body: AISI SS 304
8. Trays : Minimum 6 Removable Trays for easy handling
9. Chart Recorder ; Weekly circular Chart Recorder
10. Should have visual, audible indication for door open, high and low temperature and power on.
11. Should have a battery backup of at least 2 hrs for indicators and alarms.
12. Temperature recording chart (300nos) and ink pen (5 nos) shall be supplied free of cost.
13. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.

