



Running Contract Details	
Equipment Name	Puretone Audiometer
Running Contract Valid Till	06-12-2020
Tender Ref No	KMSCL/EP/T288/198A/2018(R)
Tendered Quantity	200
Supplier Name	M/s Auditivo Hearing Services Private Limited
GST No	07AALCA7725L1Z8
Model & Make	AUDITIVO MICRA+ WITH TDH39 / AUDITIVO HEARING
Unit rate (Rs)	58,500.00
CGST 0.00%	0.00
SGST 0.00%	0.00
IGST 5.00%	2,925.00
Total Cost(Exl.KMSCL S.C) Rs.	61,425.00
Flood Cess (1%) Rs.	0.00
Service Charges 7% + GST 18%	4,832.10 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	66,257.10
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
309 Oriental House 19 Community Centre Yusuf Sarai New Delhi	Contact Person	Ajay Kuamr Mathur
	Phone	1149874415
	Mobile No	9810112042
	Email	auditivohearing@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Puretone Audiometer	61425	4832.1	66257.1
		61425	4832.1	66257.1

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Puretone Audiometer							
Labour	1,000.00	1,500.00	2,000.00	0.00	0.00	0.00	0.00
Comprehensive	2,500.00	3,500.00	4,500.00	0.00	0.00	0.00	0.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Puretone Audiometer

1. Transducer: headset with earphones (TDH 39/TDH 39P/TDH 49) & supra aural ear cushions (MX41 AR), insert earphones, bone vibrator (Radio ear b-71), Talk over microphone, Talk back microphone.
2. Patient response switch.
3. Patch wards for two room situation
4. Monitoring head set
5. Signal level should be indicated by VU meter (speech audiometry).
6. Options for external input.
7. Stimuli Type: Pure tones (continuous, pulse tone, interruptor), Narrowband noise, speech noise, speech (monitored live voice and recorded voice)
8. Frequencies: 250Hz to 8000Hz in octaves for air conduction 250Hz to 4000Hz in octaves for bone conduction.
9. Intensity:
 - a. 10dB HL to 120dB HL in steps of 5dB for air conduction
 - b. 10dB HL to 70 dB HL in steps of 5dB for bone conduction from 250Hz to 4000Hz
10. Power supply: UPS/Battery backup & mains.
11. Self-calibration

12. Power supply: 220 Volts, 50Hz