



Running Contract Details	
Equipment Name	Sound proof system
Running Contract Valid Till	03-02-2021
Tender Ref No	KMSCL/EP/T296/679/2018(R)
Tendered Quantity	25
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Model & Make	Audiometry Booth â€“30 STC / Sinchana Associates
Unit rate (Rs)	15,35,000.00
CGST 9.00%	1,38,150.00
SGST 9.00%	1,38,150.00
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	18,11,300.00
Flood Cess (1%) Rs.	15,350.00
Service Charges 7% + GST 18%	1,26,791.00 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	19,53,441.00
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgical.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Sound proof system	1826650	126791	1953441
		1826650	126791	1953441

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Sound proof system

1. A single chamber sound treated room.
2. The wall of the chamber will be cavity walls filled with Sound insulation material.
3. The inside of the cubicle will have acoustic treatment on wall and ceiling and sound absorbent flooring.
4. Inner size of the cubicle should not be less than 8 ft. X 8 ft.
5. Air Conditioner: Split type at least 1 ton. A.C. should be fitted inside the chamber.
6. Door: There will be one sound treated door for the main entry which will be of 750mm (W).
7. Chamber should have one double glass window of at least 2ft.X 2ft.
8. Total internal electrical wiring including required electrical points, fixing of tube and fixtures for proper intensity in lux shall be carried out.
9. Equipment connections including patch panel and audio sockets at desired points should be provided.
10. The finished room must satisfy
 - a. Noise criteria shall be 30dB(A) or better
 - b. Transmission loss 30dB(A) or better
 - c. Reverberation 0.6 sec or better