



Running Contract Details	
Equipment Name	Sterile connecting device
Running Contract Valid Till	27-09-2019
Tender Ref No	KMSCL/EP/T230/1029/2017
Tendered Quantity	4
Supplier Name	M/s Haemolab Equipments
GST No	32AAHFH4793A1ZN
Model & Make	Compodoc / Fresenius Kabi Germany
Unit rate (Rs)	7,50,000.00
CGST 6.00%	45,000.00
SGST 6.00%	45,000.00
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	8,40,000.00
Flood Cess (1%) Rs.	7,500.00
Service Charges 7% + GST 18%	61,950.00 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	9,09,450.00
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.64/3652 (Old 32/2605 â€“ D) - Nilavareth Lane Azad Road - Kaloore Cochin - 682017	Contact Person	Achutha Prakash AJ
	Phone	0484 6999966
	Mobile No	90371 77781, 09387645876
	Email	prakashachutha@gmail.com, sales@haemolab.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Sterile connecting device	847500	61950	909450
		847500	61950	909450

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Sterile connecting device							
Labour	32,745.00	32,745.00	32,745.00	32,745.00	32,745.00	32,745.00	32,745.00
Comprehensive	72,570.00	72,570.00	72,570.00	72,570.00	72,570.00	72,570.00	72,570.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Sterile connecting device

1. Should accommodate and weld all types of blood bags tubing in use in our country
2. The welding should be seamless
3. Should be capable of joining wet-wet/wet-dry/dry-dry tubes.
4. Welding should not affect the quality of the tube in terms of its physical and chemical properties and it should not cause hemolysis.
5. It should have LED indicators to display the actual status of the ongoing procedural steps and audio- visual alarm system for any functional irregularities.
6. Requirement for tube length to be welding/docking should be as small as possible
7. The welding accessories should be available with the local agent throughout year.
8. The consumable cost per 100 pieces/100 welds will be taken into account during price evaluation.
9. Certifications:
 - a. CE class II A or US FDA certified.
 - b. Quality certifications: ISO 13485 certified
10. Should be supplied with suitable UPS with 30 minutes backup along with the equipment.