



Running Contract Details	
Equipment Name	Trop T/I Analyzer
Running Contract Valid Till	17-01-2021
Tender Ref No	KMSCL/EP/T294/239/2018
Tendered Quantity	104
Supplier Name	M/s Roche Diagnostics India Pvt Ltd
GST No	33AAACA0266H1Z7
Model & Make	Roche Cobas h232 / Roche Diagnostics GmbH
Unit rate (Rs)	1,25,000.00
CGST 9.00%	11,250.00
SGST 9.00%	11,250.00
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	1,47,500.00
Flood Cess (1%) Rs.	1,250.00
Service Charges 7% + GST 18%	10,325.00 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	1,59,075.00
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Sreyas Chamiers Tower 8th Floor West wing New No. 37 Old NO 23&24 Chamiers Road Teynampet Chennai-600018	Contact Person	Deepak Ranjan
	Phone	044-43900300
	Mobile No	919742746564
	Email	t.manjunathan@roche.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Trop T/I Analyzer	148750	10325	159075
2	Test Cards	463.3	33.87	497.17
		149213.3	10358.87	159572.17

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Trop T/I Analyzer							
Labour	8,750.00	8,750.00	8,750.00	8,750.00	8,750.00	8,750.00	8,750.00
Comprehensive	22,000.00	22,000.00	22,000.00	22,000.00	22,000.00	22,000.00	22,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Trop T/I Analyzer

1. Should be able to measure high sensitive quantitative measurement of Troponin-T / Troponin I. Should be at least third generation assay.
2. It should be a portable unit and have inbuilt battery backup.
3. It should have parameter wise QC checking facility.
4. Should have a memory capacity of minimum 500 test results with date time and comments.
5. Should have minimum 200 code chip records (100 test+100QC).
6. Should have a minimum sample volume of approximately 300 micro litre.
7. Should have an operating temperature ranging from 18-30 degree Celsius.
8. Should work on 200-240Vac 50Hz power supply.
9. It should have an internal printing facility or supplied with an external thermal printer.
10. Should be supplied with 50 strips of Troponin T/I Parameter.
11. Should be provided with calibration certificate issued by the manufacturer at the time of installation.
12. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL.