



Running Contract Details	
Equipment Name	Twin Phantom Head Unit
Running Contract Valid Till	12-09-2019
Tender Ref No	KMSCL/EP/T231/949/2017
Tendered Quantity	35
Supplier Name	M/s Confident Dental Equipments Pvt Ltd
GST No	32AAACC9198C1ZS
Model & Make	C-125A/Confident
Unit rate (Rs)	85,140.00
CGST 6.00%	5,108.40
SGST 6.00%	5,108.40
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	95,356.80
Flood Cess (1%) Rs.	851.40
Service Charges 7% + GST 18%	7,032.56 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	1,03,240.76
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
No.4 - 1st Floor Shivdas Tower Opp. Maymoon Theatre Chittoor Road Cochin - 11	Contact Person	Mr.sarath,Mini
	Phone	
	Mobile No	9995592737,8089383364
	Email	cdelcochin@vsnl.net,marketingmngr@confidental.org,surgident@vsnl.net

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Twin Phantom Head Unit</b>	96208.2	7032.56	103240.76
		<b>96208.2</b>	<b>7032.56</b>	<b>103240.76</b>

## **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Twin Phantom Head Unit**

1. Should have Phantom head table with two students working place
2. The Top surface should be made completely out of stainless steel
3. Should have LED operating light
4. Should have detachable, autoclavable antiglare should help in avoiding composite curing
5. Should have 30,000 Lux at 80 cm distance
6. Should have Reflector type, light should come from the reflector not direct focusing by LED
7. Should have Intensity control & on/off should be smooth with the non touch sensor switch
8. The light should have 3 axis movements
9. Light arm should be aluminium extruded. Well balanced up and down movement of 575 mm to 600 mm
10. Modular fitted with : Airotor control box fitted with transparent water tank, tubing with nut
11. Phantom head with mankin body & pneumatic piston for all the movement of neck, TMJ movement, even provision to fix TMJ in open position
12. Heavy duty phantom head with the excellent quality face mask, total closed oral cavity, drain nipple for drain of water
13. Working stool : Five wheel surgical stool with, pneumatic piston for height adjustment, seat and back rest