



Running Contract Details	
Equipment Name	Ultrasonic Scaler
Running Contract Valid Till	02-05-2020
Tender Ref No	KMSCL/EP/T257/400C/2017
Tendered Quantity	200
Supplier Name	M/s Marvel Agencies
GST No	32AJZPA3421F1ZJ
Model & Make	P.7L / Medical Baolai
Unit rate (Rs)	43,304.00
CGST 6.00%	2,598.24
SGST 6.00%	2,598.24
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	48,500.48
Flood Cess (1%) Rs.	433.04
Service Charges 7% + GST 18%	3,576.91 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	52,510.43
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
17/1682 D1 Thayyil Arcade Opp. Coronation Theatre Kozhikode -673004 Kerala	Contact Person	
	Phone	
	Mobile No	
	Email	marvelagenciesltd@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Ultrasonic Scaler</b>	48933.52	3576.91	52510.43
		<b>48933.52</b>	<b>3576.91</b>	<b>52510.43</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year
<b>Ultrasonic Scaler</b>			
Labour	0.00	0.00	0.00
Comprehensive	0.00	0.00	0.00

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Ultrasonic Scaler**

1. Should be a Piezoelectric scaler
2. Should have a detachable Autoclavable hand piece with 5 tips
3. Should have a High power turbo mode and a low power perio mode.
4. Should have removable hose with liquid flow control knob.
5. Should have a torque tool for tightening of the tip.
6. Should have a tip adapter
7. Should have automatic power feedback control
8. Should have water pressure 25-60 psi.
9. Should have Flow Rate Minimum setting (CCW) < 15 ml/min & Maximum setting (CW) > 55 ml/min
10. Should have a basic vibration frequency of 50 KHZ
11. Should be supplied with an additional hand piece & 5 tips
12. Should have foot control operation.(Base metallic type)
13. Ultrasonic scalar stand should be provided