



Running Contract Details	
Equipment Name	VDRL rotator
Running Contract Valid Till	12-02-2021
Tender Ref No	KMSCL/EP/T304/134/2018(R)
Tendered Quantity	500
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Model & Make	PVD-R01 / Rotek Plus
Unit rate (Rs)	5,850.00
CGST 9.00%	526.50
SGST 9.00%	526.50
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	6,903.00
Flood Cess (1%) Rs.	58.50
Service Charges 7% + GST 18%	483.21 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	7,444.71
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>VDRL rotator</b>	6961.5	483.21	7444.71
		<b>6961.5</b>	<b>483.21</b>	<b>7444.71</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year
<b>VDRL rotator</b>			
Labour	1,100.00	1,100.00	1,100.00
Comprehensive	500.00	500.00	500.00

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :VDRL rotator**

1. Should be a table top model with platform size 300x300mm.
2. Should be suitable for wide variety of rotating and mixing applications.
3. Should have built-in 0 to 30 minutes digital timer.
4. Should have a maximum speed of 180 RPM and suitable for serological, VDRL tests and other tuned clinical diagnostic procedures.
5. Should have spring holders for flasks and blood bottles.
6. Should work with input 200 to 240Vac 50 Hz supply.