



Running Contract Details	
Equipment Name	Vein Detecting Transilluminator/Device
Running Contract Valid Till	22-03-2021
Tender Ref No	KMSCL/EP/T307/1217/2019(R)
Tendered Quantity	50
Supplier Name	M/s Sandor Mediccaids Private Limited
GST No	
Model & Make	Veinus Pro / Vphore Labs
Unit rate (Rs)	80,000.00
CGST 0.00%	0.00
SGST 0.00%	0.00
IGST 12.00%	9,600.00
Total Cost(Exl.KMSCL S.C) Rs.	89,600.00
Flood Cess (1%) Rs.	800.00
Service Charges 7% + GST 18%	6,608.00 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	97,008.00
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
8-2-326/5 Road No 3 Banjara Hills Hyderabad - 500034 Telangana	Contact Person	M.R Suresh Kumar
	Phone	040-23357048
	Mobile No	9848022527
	Email	suresh@sandor.co.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Vein Detecting Transilluminator/Device	90400	6608	97008
		90400	6608	97008

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3rd Year	4th Year	5th Year
Vein Detecting Transilluminator/Device			
Labour	3,200.00	3,520.00	3,872.00
Comprehensive	5,600.00	6,160.00	6,776.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Vein Detecting Transilluminator/Device

1. Should be a lightweight portable and handheld non contact vein finder/viewer
2. Should use Near Infra Red (NIR) light source.
3. Should have multiple modes for viewing, based on different skin tones and body characteristics of patients
4. Should be able to view veins at 8mm depth
5. Should have rechargeable battery which will last at least 4 hours of backup
6. Should be ideal for adults and infants.
7. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid