



Running Contract Details	
Equipment Name	Video Intubating Bronchoscope
Running Contract Valid Till	22-01-2021
Tender Ref No	KMSCL/EP/T298/1207/2018(R)
Tendered Quantity	51
Supplier Name	M/s Lakshmi Surgicals
GST No	32AACFL2603A1ZB
Model & Make	Karl Storz Video Intubating Bronchoscope / Karl Storz GmbH & Co Germany
Unit rate (Rs)	18,21,445.76
CGST 0.00%	0.00
SGST 0.00%	0.00
IGST 0.00%	0.00
Total Cost(Exl.KMSCL S.C) Rs.	18,21,445.76
Flood Cess (1%) Rs.	0.00
Service Charges 7% + GST 18%	1,50,451.42 (Applicable for purchase through KMSCL only)
Total Cost(Incl. KMSCL S.C) Rs.	19,71,897.18
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
203XIX -City Towers Bank Jn -Market Road Aluva -Ernakulam - 683101 Kerala	Contact Person	Mr. Praveen
	Phone	
	Mobile No	9995271784
	Email	lakshmisurgicals2005@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Video Intubating Bronchoscope	1821445.76	150451.42	1971897.18
2	Monitor	849779.95	58984.73	908764.67
		2671225.71	209436.15	2880661.85

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Video Intubating Bronchoscope							
Labour	27,242.40	29,966.64	32,963.30	36,259.64	39,885.60	43,874.17	48,261.59
Comprehensive	54,484.81	59,933.29	65,926.60	72,519.26	79,771.21	87,748.34	96,522.82

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Video Intubating Bronchoscope

1. Should supply with CCD/CMOS Technology
2. The processor should be on high definition platform of resolution 1920 x 1080 P
3. Should possess LED light source equivalent to 150 W xenon or more and life of 10,000 hours or more
4. Should have high definition video images, recording should be available via internal or external methods
5. Deflection up/down : 1200/1200or more
6. Direction of view : 00
7. Angle of view : 900 – 1200
8. Working length : 600 mm or above
9. Working channel inner diameter : 1.2 – 2 mm
10. Distal tip outer diameter : 4 – 5.5mm
11. Should have option for suction valve
12. Should be splash-proof according to IP54 & suitable for wipe disinfection
13. Should be US FDA and CE approved

A. Monitor – medical grade, LCD panel with LED backlit -1 no

1. Should have screen size 21 inch or above with 1920 x 1080 pixel resolution
2. Should have a camera input, a USB and a HDMI port
3. Should have optimized user interface, video and image capturing in real time on internal hard disk in which the data is transferable, playback of recorded video clips and still images & data transfer from SD card to USB flash drive

B. Compact multi-rack trolley with wheel lock system manufactured by the OEM Company -1no (to hold the above instrument and all accessories)

1. It should have at least 6 power sockets (16/6amps combination socket) to connect the processor, monitor etc. The power sockets should be universal type accepting Indian/European/US model plugs
2. It should also have a stand to hold the scope in between or after the procedure.

C. Online UPS of suitable capacity with minimum 30 minutes backup.

D. Scope storage box – 1no

E. Scope sterilization tray -1no

F. Should supply with the following

1. Leakage Tester
2. Tube Holder
3. Cleaning Brush
4. Irrigation Adaptor suction Valve
5. Bronchoscope Insertion tube
6. Size 4 plug, black
7. Package of 10 protection Cap

