



# KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/216/2025

Dated: 19.11.2025

## NOTICE

**Sub:-** Tender No.T563/ 2025 - Results of demonstration/ document evaluation & date of price bid opening –  
Diathermy Surgical- reg:-

- Ref:-**
1. Tender Inviting Notice published dated 05.03.2025
  2. Technical bid opening held on 24.01.2025 @ 11.00 AM
  3. Eligible list published on 18.06.2025
  4. Demonstration/ document evaluation of Diathermy Surgical conducted on 17.09.2025

I. The reports of technical document evaluation/ demonstration of Diathermy Surgical as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Diathermy Surgical	M/s. UNIVERSAL AGENCIES	EPSILON-400V2/CUADRA MEDICAL
		M/s. JBS Meditec India P Ltd	E -Lite/Alan Electronic
		M/s. ANTS LIFECARE PRIVATE LIMITED	Easy +/-Xcellance

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

### 1. M/s.BIOCARE DIAGNOSTICS

Rejected model: ENSURG300/ENERTECH

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
5	Should have return electrode contact safety.	Not meeting specification

### 2. M/s. C CUBE ADVANCED TECHNOLOGIES

Rejected model: SC-D300/C CUBE

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
12	Should be supplied with reusable flexible silicon rubber patient return plate with return electrode safety 1 No.	Not meeting specification



3. M/s. Divlabs Healthcare Pvt. Ltd

Rejected model: Trinity/ Divlabs

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
12	Should be supplied with reusable flexible silicon rubber patient return plate with return electrode safety 1 No.	Not meeting specification

III. The online price bid of the eligible tenderers mentioned above will be opened on <sup>21</sup>.11.2025 @ 11.AM at the etender portal.

**Dr Arun P V**  
General Manager