

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/10408/2024

Dated: 05.04.2025

NOTICE

Sub:- Tender No.T544(R)/ 2024 - Results of demonstration/ document evaluation & Date of pricebid opening – various equipment reg;-

- Ref:-**
1. Tender Inviting Notice published dated 11.06.2024
 2. Technical bid opening held on 05.07.2024 @ 11.00 A M
 3. Eligible list for demonstration published dated 15.01.2025
 4. Demonstation/ document evaluation of various equipment conducted on 31.01.2025

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1	Pulse Oxymeter Baby And Adult	M/s. TRANS HEALTH CARE INDIA PRIVATE LIMITED	Pulsat 9/Trans healthcare
		M/s. Biomedical Engineering Company	RAD 97/Masimo Corporation
		M/s. BIOCARE DIAGNOSTICS	MD2000A/Beijing Choice Electronic
2.	Infantometer	M/s. Kerala Surgical Equipment Co.	Krups/Doctor Beliram & Sons
		M/s. LAKSHMI AGENCY	LAKSHMI AGENCY
3.	Weighing Machine Baby (Electronic)	M/s. Kerala Surgical Equipment Co.	BWS 2121/Pragathi Medical
		M/s. Marc Weighing Systems	MRT.20 BABY SCALE/ MARC WEIGHING SYSTEM
		M/s. Universal Agencies	MW-20K/MEDI WAVES
4.	Salters Baby Weighing Machine	M/s. Kerala Surgical Equipment Co.	Doctor Beliram & Sons

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. KMSCL/EP/T544(R)/105B/2024: Pulse Oxymeter Baby And Adult

1. M/s. Skanray Technologies Private Limited

Rejected model: M/s. Skanray Technologies Private Limited

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
15	Should have perfusion index numerical display in the measurement range of at least 0.02 to 20.	Not available

2. M/s. MITI ELECTRONICS PVT LTD

Rejected model: BT 720/ Bistos Co Ltd

Reason for rejection: Sample not produced

B. KMSCL/EP/T544(R)/145B/2024: Salters Baby Weighing Machine

1. M/s. Marc Weighing Systems

Rejected model: 235-6M/MARCWEIGHINGSYSTEM

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
3	Graduation : 100 gms	500 gms
5	Should be supplied with top and bottom hook with weighing trouser	No weighing trouser

III. The online price bid of the eligible tenderer mentioned above will be opened on 09.04.2025 @ 11.AM at the e-tender portal.

Dr. Shibulal A
General Manager