



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id :eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

Equipment Division

www.kmscl.kerala.gov.in

No. I/378270/2026

Date:10-09-2026

Notice

Sub:- Tender No.T574(R)/ 2025 - Results of demonstration/ document evaluation & Date of pricebid opening – Self-Retaining Brain Retractor System - reg.

Ref:- 1. Tender Inviting Notice published dated 14.05.2025
2. Technical bid opening held on 28.06.2025 @ 11.00 A M
3. Eligible list for demonstration published dated 24.09.2025
4.Demonstration/ document evaluation of Self-Retaining Brain Retractor System is conducted on 16.12.2025 & 06.01.2026

I. The reports of technical document evaluation/ demonstration of Self-Retaining Brain Retractor System as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models/ OEM
1.	Self-Retaining Brain Retractor System	M/s SOUTH INDIA SURGICAL CO. LTD	DORO HEADREST SYSTEM , SISCO LEYLA Retractor/ PMI GmBH, Germany

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s Aswini Health Specialities India Pvt Ltd
Rejected model; LEYLA Retractor/B Braun-Aesculap
Reason for rejection: Sample not produced

B. M/s STAAN BIO-MED ENGINEERING PRIVATE LIMITED
Rejected model: 3-pin Fixator with Leyla Retractor/STAAN BIOMED
Reason for rejection:

Clause No	Technical Specification	Deviations Noted
7	Star bursts for attaching the neuro navigation adaptor should present in the side of the skull clamp	Not meeting the specification

III. The online price bid of the eligible tenderer mentioned above will be opened on 14.09.2026 @ 11.AM at the e-tender

Rakesh R U
GENERAL MANAGER