



EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/10357/2025

Dated: 29.05.2026

NOTICE

Sub:- Tender No.T578/ 2025 - Results of demonstration/ document evaluation & Date of pricebid opening – Reagent Refrigerator - Various Types- reg.

- Ref:-**
1. Tender Inviting Notice published dated 28.07.2025
 2. Technical bid opening held on 21.08.2025 @ 11.00 A M
 3. Eligible list for demonstration published dated 08.11.2025
 4. Reports of Demonstration/ document evaluation of Reagent Refrigerator - Various Types is received on 16.12.2025 & 06.01.2026

I. The reports of technical document evaluation/ demonstration of Reagent Refrigerator - Various Types as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models/ OEM
1.	Reagent Refrigerator - Various Types	M/s. QUBEC BIOSOLUTIONS	GLR 300AC ,GLR 500AC
		PRIVATE LIMITED	,GLR 800A/ Godrej
		M/s. New India Surgicals and Chemicals	YC-315LG & YC-52OLG / Blue star

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

1. M/s. BIOLINX INDIA
Rejected model: Biocool 400, Biocool 500 Biocool 700/ Biolinx India P
Reason for rejection: Sample not provided.
2. M/s. New India Surgicals and Chemicals
Rejected model: FRC.V-1000/ Cryo Scientific for 750-800 L
Reason for rejection: Sample not produced
3. M/s. Mark Enterprises
Rejected model: MLR-03 / MLR-06/: Mark Enterprises
Reason for rejection:

Clause No	Technical Specification	Deviations Noted
5	Temperature should be adjustable from 2 to 8 degree.	Not meeting the specification

4. Labline Instruments
Rejected model: LMR 14 ,LMR 21 ,LMR 28/Labline Instruments
Reason for rejection: Sample not produced

III. The online price bid of the eligible tenderer mentioned above will be opened on ⁰².05.2026 @ 11.AM at the e-tender portal.

Dr Arun P V
General Manager