



Running Contract Details	
Equipment Name	Micro Centrifuge
Running Contract Valid Till	27-05-2027
Tender Ref No	KMSCL/EP/T550/583A/2024
Tendered Quantity	10
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Micro Centrifuge <i>Model & Make : RM-03 PLUS with 12x1.5ml/2.0ml & PCR Strip Rotor (2x8x0.2ml) / REMI ELEKTROTECHNIK LTD</i>	61348.2 Incl.GST :18%	4294.37	65642.57
		61348.2	4294.37	65642.57
Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Micro Centrifuge				
Labour	1,000.00	1,000.00	1,000.00	
Comprehensive	2,000.00	2,500.00	3,000.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Micro Centrifuge

Technical Specification

Equipment: Micro Centrifuge

1. The equipment should be a bench-top high speed centrifuge.
2. Should be able to reach RCF of 16,000g.
3. Should be microprocessor controlled with multiple user definable programmes.
4. Should be able to program RPM, RCF, acceleration and deceleration.
5. Automatic lid lock should be available.
6. Noise level should be below 60dB.
7. Should be supplied with a fixed angle rotor (12 x 1.5/ 2ml) with RCF 16,000 g or above.
8. Should work with 220v – 230v, 50Hz AC supply
9. Rapid acceleration and deceleration for quick spins.
10. Dedicated button for quick spin function.