



# KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/7790/2025

Dated: 17.12.2025

## NOTICE

**Sub:-** Tender No.T573/ 2025 - Results of demonstration/ document evaluation & Date of pricebid opening –  
Laminar Flow Cabinet- reg.

- Ref:-**
1. Tender Inviting Notice published dated 14.05.2025
  2. Technical bid opening held on 02.06.2024 @ 11.00 A M
  3. Eligible list for demonstration published dated 27.08.2025
  4. Demonstation/ document evaluation of Laminar Flow Cabinet equipment is conducted on 06.11.2025

I. The reports of technical document evaluation/ demonstration of Laminar Flow Cabinet as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Model
1	Laminar Flow Cabinet	M/s Labline Instruments	VLF-4/Labline

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s Zenotrex Medico Pvt. Ltd

Rejected model: ZPL-L01F/ Zenotrex Medico

Reason for rejection: Sample not produced

B. M/s Yorco Sales Pvt Ltd

Rejected model: YSL/YSI-189/ Yorco Sales

Reason for rejection: Sample not produced

C. M/s Instromedix India Pvt. Ltd

Rejected model: PLT-271/ P.L TANGO

Reason for rejection;

Clause No	Technical Specification	Deviations Noted
2	Work surface should be SS304.	Not meeting the specification
	Comments	Work surface side panels not stainless steel

D. M/s UNIVERSAL AGENCIES

Rejected model; YSI-189/YORK

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
3	Should have M.S body with sufficient protective coating.	Not meeting the specification
	Comments	Sliding glass not covered fully which makes cleaning difficult leads to bacterial contamination in laboratory environment

E. M/s Dinesh Scientific

Rejected model; DS-LAF-4X2X2/Dinesh

Reason for rejection: Sample not produced

III. The online price bid of the eligible tenderer mentioned above will be opened on 2112.2025 @ 11.AM at the e-tender portal.

Dr Arun P V  
General Manager