



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/10357/2025

Dated: 19.05.2026

NOTICE

Sub:- Tender No.T578/ 2025 - Results of demonstration/ document evaluation & Date of pricebid opening – various equipment- reg.

Ref:- 1. Tender Inviting Notice published dated 28.07.2025

2. Technical bid opening held on 16.09.2025,10.09.2025 & 21.08.2025 11.00 A M

3. Eligible list for demonstration published dated 08.11.2025

4. Demonstration/ document evaluation of various equipment conducted on 12.01.2026

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Model
1	Video Laryngoscope Model B	M/s. Arkon Medical Systems	VL3D/Shenzhen HugeMed
		M/s. LAKSHMI AGENCY	S-VL-O1/Sensorendo
2.	Ventilator Portable	M/s. AIR LIQUIDE MEDICAL SYSTEMS PVT LTD	MONNAL T60/Air Liquid
		M/s. Sree Gokulam Healthcare Pvt. Ltd	SV300/Shenzhen Mindray

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. KMSCL/EP/T578/199B/2025: Video Laryngoscope Model B

1. M/s. Hansraj Nayyar Mediquip

Rejected model: AVL-2P-6360/ Anaesthetics

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
7	Should have passed the drop test for one meters	Not provided
9	Should have facility to run independently on a battery and back up should have minimum four hours. The rate for the battery shall be offered in the BOQ and the same will be fixed for 5 years from the date of price bid opening. The rate will be taken for evaluation.	Only two hours
10	The system should be supplied with a set of different sizes of disposable blade size 1,2,3,4 and one additional blade for difficult intubation	Not meeting the specification.

11	disinfection (without battery)	Not meeting the specification
12	Should supply the following blades free of cost along with the machine - 25 nos. of size 1, size 2, size 3, size 4 and 10 numbers of 'additional blade for difficult intubation'	Not meeting the specification

III. The online price bid of the eligible tenderer mentioned above will be opened on ~~03.05.2026~~ ^{25.05.2026} @ 11.AM at the e-tender portal.

Dr Arun P V
General Manager