



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/15170/2024

Dated: 30.04.2025

NOTICE

Sub:- Tender No.T547/ 2024 - Results of demonstration/ document evaluation & Date of pricebid opening – various items - reg.

- Ref:-**
1. Tender Inviting Notice published dated 27.08.2024
 2. Technical bid opening held on 03.10.2024 & 23.09.2024 @ 11.00 A M
 3. Eligible list for demonstration published dated 21.12.2024
 4. Demonstation/ document evaluation of various items conducted on 10.04.2025

I. The reports of technical document evaluation/ demonstration of various items as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

| SI No | Equipment Name | Selected Bidders | Selected model |
|-------|----------------|-----------------------------------|--|
| 1. | Blood Lancets | M/s. UNIVERSAL AGENCIES | SWISS MED/GLOBAL SURGIMED |
| | | M/s. LAKSHMI AGENCY | Recombigen |
| 2. | Micro Slides | M/s. UNIVERSAL AGENCIES | BLUESTAR PLAIN MICROSLIDES/POLAR INDUSTRIAL |
| | | M/s. PLAZA INDUSTRIAL CORPORATION | PLAZA/PLAZA INDUSTRIAL |
| | | M/s.UNIQUE SCIENTIFIC | Microslide 7101 with Tissue paper, vacuum packed |
| 3. | Chloroscope | M/s. UNIVERSAL AGENCIES | LTEK CHLOROSCOPE/ LTEK SYSTEMS |
| | | M/s. LAKSHMI AGENCY | Rakiro Biotech Systems |

II. The other tenders participated in the tender are rejected due to the reasons mentioned against their names below;

A. KMSCL/EP/T547/736/2024: Blood Lancets

1. M/s. AMKAY PRODUCTS PVT. LTD

Rejected model: AMKAY PRODUCTS

Reasons for rejection:

| Clause No | Technical Specification | Deviations Noted |
|-----------|---|---------------------|
| 7 | Material: Chrome-polished stainless steel | Not stainless steel |
| 9 | Initial sterilisation method: Ethylene oxide gas or Gamma radiation | Not mentioned |

2. M/s. Bio-chem Engineering

Rejected model: Twist Lancets

Reason for rejection: Sample not produced

B. KMSCL/EP/T547/737/2024: Micro Slides

1. M/s. BIOCRAFT SCIENTIFIC SYSTEMS PVT. LTD

Rejected model: BIA-1/Bio craft scientific

Reason for rejection:

| Clause No | Technical Specification | Deviations Noted |
|-----------|---|------------------|
| 5 | Edge quality: The edge of the slides shall be ground so that they are free from sharp edges and chips | Irregular edges |
| 7 | Slides should be individually interleaved with tissue paper | Not available |

III. The online price bid of the eligible tenderer mentioned above will be opened on 02.05.2025 @ 11.AM at the etender portal.

Dr. Shibulal A
General Manager