



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/10357/2025

Dated: 13.03.2026

NOTICE

Sub:- Tender No.T569/ 2025 - Results of demonstration/ document evaluation & Date of pricebid opening – various equipment- reg.

Ref:- 1. Tender Inviting Notice published dated 19.03.2025

2. Technical bid opening held on 07.04.2024 @ 11.00 A M

3. Eligible list for demonstration published dated 27.08.2025

4. Demonstration/ document evaluation of various equipment conducted on 08.01.2026

I. The reports of technical document evaluation/ demonstration of vrious equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Model
1	BINOCULAR MICROSCOPE – Type I	M/s. UNIVERSAL AGENCIES	LABOMED LX 400/N.K.JAIN
2	BINOCULAR MICROSCOPE – Type II	M/s. UNIVERSAL AGENCIES	LABOMED LX 500/N.K.JAIN
3.	Blood Collection Chair	M/s. LabX	FLABO-X/ Remi

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. KMSCL/EP/T578/13A/2025: Binocular Microscope - LED Various Types

1. BINOCULAR MICROSCOPE – Type I

a. M/s. iLABS

Rejected model: iscope CLX-I / iLABS

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
2	Eyepieces - Highest quality 10 X wide field anti fungus field eyepiece, FOV 20mm.	Not meeting the specification

b. M/s. Mark Enterprises

Rejected model; MMB-02/: Mark Enterprises

Reason for rejection: Sample not produced

c. M/s. Scientific Medical Equipment (Kerala) Pvt.Ltd

Rejected model; Binocular Microscope Type I –kxi 2000

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
-----------	-------------------------	------------------

12	competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid	Not a valid certificate
----	--	-------------------------

d. M/s. Micron Instrument Industries

Rejected model; MAKE MICRON OPTIK CRESTA XS 50/ Micron Instrument

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
12	Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid	Not a valid certificate

e. M/s. UNIQUE SCIENTIFIC

Rejected model: RH-85UXL/UNILAB

Reason for rejection: Sample not produced

f. M/s. Sudheer Scientific Works

Rejected model; SKC-CXLed , /Sudheer Scientific

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
12	Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid	Not a valid certificate

g. M/s. Hoverlabs

Rejected model; HVSBM20 , /HOVERLABS

Reason for rejection: Sample not produced.

2. BINOCULAR MICROSCOPE – Type II

a. M/s. iLABS

Rejected model; iscope XXB / iLABS

Clause No	Technical Specification	Deviations Noted
2	Eyepieces - Highest quality 10 X wide field anti fungus field eyepiece, FOV 20mm.	Not meeting the specification

b. M/s. Mark Enterprises

Rejected model: MMB-02/ Mark Enterprises

Reason for rejection: Sample not produced

c. M/s. Scientific Medical Equipments (Kerala) Pvt.Ltd

Rejected model: Binocular Microscope Type II - c3/LABEX

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
12	Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid	Not a valid certificate

d. M/s. Micron Instrument Industries

Rejected model; MAKE MICRON OPTIK CRESTA ZS 501/ Micron Instrument

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
12	Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid	Not a valid certificate

e. M/s. UNIQUE SCIENTIFIC

Rejected model; RH-85UXL/UNILAB

Reason for rejection: Sample not produced

f. M/s. Sudheer Scientific Works

Rejected model; SKC-440 IR/Sudheer Scientific

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
-----------	-------------------------	------------------

12	competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid	Not a valid certificate
----	--	-------------------------

g. M/s. Hoverlabs

Rejected model; HVSBM22/HOVLRLABS

Reason for rejection: Sample not produced.

B. KMSCL/EP/T578/1801/2025: Blood Collection Chair

a. M/s. Mark Enterprises

Rejected model; MBDC-05/Mark Enterprises

Reason for rejection: Sample not produced

III. The online price bid of the eligible tenderer mentioned above will be opened on ¹⁹03.2026 @ 11.AM at the e-tender portal.

Dr Arun P V
General Manager