KERALA MEDICAL SERVICES CORPORATION LTD
(A Government Of Kerala Undertaking)
Thycaud.P.O., Thiruvananthapuram-14.
Tele Fax No. 0471-2945647
0471-2945646, 0471-2945600
E-mail id :epetenders.kmscl@kerala.gov.in
CIN : U24233KL2007SGC021616
GSTIN: 32AADCK4029M1ZK

Equipment Division www.kmscl.kerala.gov.in

No. I/415159/2026 Date:26-10-2026

Notice

Sub:- Tender No.T568(R)/ 2025 - Results of demonstration/ document evaluation & Date of pricebid opening – EMG Machine - reg.

Ref:- 1. Tender Inviting Notice published dated 05.02.2025

2. Technical bid opening held on 17.03.2025 @ 11.00 A M

3. Eligible list for demonstration published dated 21.08.2025

4.Reports of demonstation/ document evaluation of EMG Machine is received on 08.10.2026

I. The reports of technical document evaluation/ demonstration of EMG Machine as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models/ OEM
1.	EMG Machine	M/s. Axis Health Care	Natus ultra Pro S100/ Natus Medical

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

1. M/s. UNIVERSAL AGENCIES
Rejected model: NS-4/MEDICAID SYSTEMS.
Reason for rejection:

Clause No	Technical Specification	Deviation Noted
1	Should be a PC based system. Should have adaptor box with dedicated keyboard on it or if possible all controls shall be on the amplifier box.	No
7	It shall have volume control ON/OFF switch on the amplifier box	No
11	It shall have volume control ON/OFF switch on the amplifier box	No

15	It shall have inbuilt battery backup for atleast 30 minutes or more.	No
	Others	
2	Should be supplied with a PC of adequate configuration having HDD of storage not less than 360 GB HDD, DVD/CD writer, Colour Printer & USB Port.	No
3	Monitors provided along with PC should be 17" LCD / TFT and Colour Printer should be Colour Inkjet Printer.	No
4	Should supply online UPS of sufficient capacity with 1 hour backup to connect all the equipments supplied except Tread Mill system.	No
6.g	Single fiber EMG facility	No
6.h	Autonomic nervous system testing kit.	No
6.i	Collision technique.	No

III. The online price bid of the eligible tenderer mentioned above will be opened on 29 .10.2026 @ 11.AM at the e-tender

Rakesh R U

GENERAL MANAGER