



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/4147/2024

Dated: 03.12.2024

NOTICE

Sub:- Tender No.T534/ 2024 - Cancellation Notice – Portable Electronic Tourniquet (Dual) - reg.

Ref:- 1. Tender Inviting Notice published dated 05.03.2024

2. Technical bid opening held on 10.05.2024 & 20.05.2024 @ 11.00 A M

3. Eligible list for demonstration published dated 28.09.2024

4. Reports of demonstation/ document evaluation of Portable Electronic Tourniquet (Dual) received on 28.11.2024

I. The reports of technical document evaluation/ demonstration of Portable Electronic Tourniquet (Dual) as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers participated in the technical evaluation are tabulated as follows:

A. KMSCL/EP/T534/125B/2024: Portable Electronic Tourniquet (Dual)

1. M/s LAKSHMI AGENCY

Rejected model: Medione/ Avrasyamed Medikal

Reason for rejection: Sample not produced

2. M/s Manchel Marketing Company

Rejected model; 210 DP+/ Shree Saraswati

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
15	Maintenance facilities like calibration, self-test, leak test should be present on the machine.	Pressure was falling on squeezing the cuff
17	Should have usb port for saving file and updates	Not available
18	Should have Fast Choice Buttons to allow to change pressure and time to the preset values, default button should allow personalized presetting of time and pressure, quick access to display features of Tourniquet Touch	Not available

III. Due to administrative reason the tender for KMSCL/EP/T534/125B/2024: Portable Electronic Tourniquet (Dual) is cancelled.

Dr. Shibulal A

General Manager

