



EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/8653/2025

Dated: 23.01.2026

NOTICE

Sub:- Tender No.T575(R)/ 2025 - Cancellation Notice – various equipment - reg.

- Ref:-**
1. Tender Inviting Notice published dated 19.03.2025
 2. Technical bid opening held on 13.06.2025 @ 11.00 A M
 3. Eligible list for demonstration published dated 21.08.2025
 4. Demonstation/ document evaluation of various equipment is conducted on 03.12.2025

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers participated in the technical evaluation are as follows:

A. KMSCL/EP/T575(R)/1209B/2025: Pediatric Bed

1. M/s. Willson Medical System Pvt. Ltd

Rejected model: WM- 5601/Willson Medical

Reason for rejection;

Clause No	Technical Specification	Deviations Noted
4	Should have Pressed steel bottom	Not meeting the specification

B. KMSCL/EP/T575(R)/1209/2025: Pediatric ICU Cot

1. M/s. Willson Medical System Pvt. Ltd

Rejected model: WM- 5122/Willson Medical

Reasonf or rejection:

Clause No	Technical Specification	Deviations Noted
11	It should be Bacteriostatic powder coating and thermosetting epoxy polyester formulated to fulfill the requirements for bacterial protection against at least 2 commonly found bacteria in Hospital environment)	Not meeting the specification

II. Due to technical reason the tender for **KMSCL/EP/T575(R)/1209B/2025: Pediatric Bed** & **KMSCL/EP/T575(R)/1209/2025: Pediatric ICU Cot** are cancelled.

Dr Arun P V

General Manager

