



# KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/10408/2024

Dated: 24.05.2025

## NOTICE

**Sub:-** Tender No.T544(R)/ 2024 -Results of demonstration/ document evaluation & Date of price bid opening – Mobile Spot Light LED Model B - reg.

**Ref:-** 1. Tender Inviting Notice published dated 11.06.2024  
2. Technical bid opening held on 15.07.2024@ 11.00 A M  
3. Eligible list published on 15.01.2025  
4. Demonstration/ document evaluation of Mobile Spot Light LED Model B is conducted on 06.03.2025 & 08.05.2025

I. The reports of technical document evaluation/ demonstration of Mobile Spot Light LED Model B as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Mobile Spot Light LED Model B	M/s. Kerala Surgical Equipment Co.	LED 100/Medik Enterprises

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s.Alan Medical & Laboratory

Rejected model: SPARX LED.O5/HARDIK MEDI TECH

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
4	It should have minimum light intensity of 40,000 lux at 1 meter	Not demonstrated
5	It should have 4300 chromaticity.	Not demonstrated
6	CRI should be minimum 90.	Not demonstrated
9	Color temperature should be 4000K-5000K.	Not demonstrated

B. M/s. Dynamic International

Rejected model: DI-005/DYNAMIC

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Comments	No model number in the light



## C. M/s.STAAN BIO-MED ENGINEERING PRIVATE LIMITED

Rejected model: EX.LED/STAAN

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
4	It should have minimum light intensity of 40,000 lux at 1 meter	20000 Lux

## D. M/s.LAKSHMI AGENCY

Rejected model: OPTILUC-11/SURGICOIN MEDIEQUIP

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
10	Stand height should be minimum 40" from the floor and Goose neck arm of length at least 65 cm	Goose neck arm of length 60 cm

## E. M/s.Thirupathi Enterprises

Rejected model: MEDFOX TULIP-B/ILLUMIX LED

Reason for rejection: Sample not produced

III. The online price bid of the eligible tenderer mentioned above will be opened on ~~21~~04.2025 @ 11.AM at the e-tender portal.

**Dr. Vinay Goyal IAS**  
Managing Director