



EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/10358/2025

Dated: 29.05.2026

NOTICE

Sub:- Tender No.T579(R)/ 2024 - Results of demonstration/ document evaluation & Date of pricebid opening – Flexible Naso-Pharygo Laryngoscope with Channel and Instruments- reg.

- Ref:-**
1. Tender Inviting Notice published dated 28.07.2025
 2. Technical bid opening held on 07.08.2025 @ 11.00 A M
 3. Eligible list for demonstration published dated 15.10.2025
 4. Demonstation/ document evaluation of Flexible Naso-Pharygo Laryngoscope with Channel and Instruments is conducted on 23.01.2026

I. The reports of technical document evaluation/ demonstration of Flexible Naso-Pharygo Laryngoscope with Channel and Instruments as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models/ OEM
1.	Flexible Naso-Pharygo Laryngoscope with Channel and Instruments	M/s Lakshmi Surgicals	Karl Storz – Rhino Laryngo Fiberscope with Telecam C3/Karl Storz

II. The other tendereres participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s Arkon Medical Systems

Rejected model: VL3S with MS-8/Shenzhen HugeMed

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
II. 1	Should have an LED light source equivalent to 150W light output.	Even though M/s. Arkon Medical Systems specifications does not match being chip on tip system
II. 2	Should have manual light intensity control.	
II. 3	Should have a cooling system.	
II. 4	Should work with input 200 to 240 Vac 50Hz supply.	
III.2	Should have one composite video outputs and one S-video output.	
III.3	Should have anti-moisture filter for fiber scopes.	
III.4	Should have fully automatic exposure control.	
III.2	Should have one composite video outputs and one S-video output.	

B. M/s Innovative Concepts

Rejected model: FI-10RBS/ PENTAX Europe

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
III. 3	Should have anti-moisture filter for fiber scopes.	Not meeting the specification
III.5	Should have automatic white balance with memory function.	Not meeting the specification

II. The online price bid of the eligible tenderer mentioned above will be opened on ⁰².05.2026 @ 11.AM at the e-tender portal.

Dr Arun P V
General Manager