



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/18239/2024

Dated: 21.10.2025

NOTICE

Sub:- Tender No.T553/ 2024 - Results of demonstration/ document evaluation & Date of price bid opening – various equipment - reg.

Ref:- 1. Tender Inviting Notice published dated 16.10.2024

2. Technical bid opening held on 28.10.2024 & 04.11.2024@ 11.00 A M

3. Eligible list for equipment demonstration published on 29.03.2025

4. Demonstration/ document evaluation of various equipment Conducted on 21.06.2025

I. The reports of technical document evaluation/ demonstration of various equipment per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Head Light for ENT OP	M/s. Lakshmi Surgicals	Karl Storz - Head Light for ENT OP/Karl Storz
		M/s Kerala Surgical Equipment Co.	GMH5/Gem Surgitech
		M/sTrivandrum Surgicals and Lab Supplies	VMS MODEL:KD/VMS
2.	Head Light for ENT OT	M/s.Lakshmi Surgicals	Karl Storz -Head Light for ENT OT/Karl Storz -Head Light for ENT OT
		M/s.Trivandrum Surgicals and Lab Supplies	VMS MODEL:KD/VMS

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. KMSCL/EP/T553/54/2024: Head Light for ENT OT

1. M/s. PIKA MEDICAL PRIVATE LIMITED

Rejected model; Bistos Head Light/Bistos(BT410F)

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
3	Should have head light adjustment side to side and up and down.	side to side adjustment not available
4	Should have focusing sleeves for uniform quality illumination.	Not available

2. M/s. UNIVERSAL AGENCIES

Rejected model: ENT-55L/UNITECH VISION

Clause No	Technical Specification	Deviations Noted
3	Should have head light adjustment side to side and up and down.	Side to side adjustment not available

3. M/s.Kerala Surgical Equipment Co

Rejected model: JW 10/Gem Surgitech

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
5	Should have an adjustable and light weight head band with lock.	Head band poor quality

III. The online price bid of the eligible tenderer mentioned above will be opened on ~~25~~10.2025 @ 11.00 a.m. at the e-tender portal.

Dr Arun P V
General Manager