



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/10401/2024

Dated: 01.04.2025

NOTICE

Sub:- Tender No.T541/ 2024 - Cancellation Notice – Ice Lined Refrigerator - reg.

Ref:- 1. Tender Inviting Notice published dated 11.06.2024

2. Technical bid opening held on 29.07.2024@ 11.00 A M

3. Eligible list published on 01.01.2025

4. Reports of demonstration/ document evaluation of Ice Lined Refrigerator received on 29.03.2025

I. The reports of technical document evaluation/ demonstration of Ice Lined Refrigerator as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers participated in the technical evaluation are tabulated as follows:

1. M /s. YORCO SALES PVT .LTD

Rejected model: YSL/YSI-950/Yorco sales

Reason for reejection:

Clause No	Technical Specification	Deviations Noted
3	Material of Construction of body: Corrosion resistant material , SS 304 (internal, external, cabinet, lid and frame)	ABS lid

2. M/s. Mark Enterprises

Rejected model: MLR-01OEM/Mark Enterprises

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
I.3	Material of Construction of body: Corrosion resistant material , SS 304 (internal, external, cabinet, lid and frame)	ABS lid, External cabinet powder coated

3. M/s. Instromedix India Pvt. Ltd

Rejected model:VLR-60/VAV instruments

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
I.3	Material of Construction of body: Corrosion resistant material , SS 304 (internal, external, cabinet, lid and frame)	Acrylic lid, External cabinet powder quoted
	Should maintain a temperature range of +2	

III.3	to + 8 degree Celsius under the ambient temperature range of +10 to +43 degree Celsius. Externally readable cabinet mounted electronic temperature display showing current temperature and alarm status. Power failure alarm and over and under temperature alarm should be provided.	No power failure alarm
III.6	Hold overtime: 48 hrs or above	Not meeting the specification

4. M/s. Eakcon systems Pvt Ltd

Rejected model: ILR 175/EAKCON

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
I.3	Material of Construction of body: Corrosion resistant material , SS 304 (internal, external, cabinet, lid and frame)	External cabinet powder quoted quoted, internal cabinet with no lid
8	Control panel and thermometer: The control panel must be positioned on the front of the unit. The on/off switch should be recessed or otherwise protected so that it is not possible inadvertently to switch the unit off.	On/Off switch not protected
III.3	Should maintain a temperature range of +2 to + 8 degree Celsius under the ambient temperature range of +10 to +43 degree Celsius. Externally readable cabinet mounted electronic temperature display showing current temperature and alarm status. Power failure alarm and over and under temperature alarm should be provided.	No power failure alarm
1	Others Additional Features : Appropriate Provision for drainage for the waste water, Optimized refrigeration system design, CFC-free.	No drainage

5. M/s. COLDCHAIN CONTROLS

Rejected model: VC-150/Coldchaincontrols

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
I. 3	Material of Construction of body: Corrosion resistant material , SS 304 (internal, external, cabinet, lid and frame)	External cabinet powder quoted
I.8	Control panel and thermometer: The control panel must be positioned on the front of the unit. The on/off switch should be recessed or otherwise protected so that it is not possible	On/Off switch not protected

	inadvertently to switch the unit off.	
III.3	Should maintain a temperature range of +2 to + 8 degree Celsius under the ambient temperature range of +10 to +43 degree Celsius. Externally readable cabinet mounted electronic temperature display showing current temperature and alarm status. Power failure alarm and over and under temperature alarm should be provided.	No power failure alarm

6. M/s. Calgon scientific CO

Rejected model: GHR 80 AC/Godrej

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
I. 3	Material of Construction of body: Corrosion resistant material , SS 304 (internal, external, cabinet, lid and frame)	Stainless steel, internal cabinet with no lid
III.3	Should maintain a temperature range of +2 to + 8 degree Celsius under the ambient temperature range of +10 to +43 degree Celsius. Externally readable cabinet mounted electronic temperature display showing current temperature and alarm status. Power failure alarm and over and under temperature alarm should be provided.	Only visual alarm during power failure
III.6	Hold overtime: 48 hrs or above	Not meeting the specification

II. Due to administrative reason the tender for **KMSCL/EP/T541/236/2024: Ice Lined Refrigerator** is cancelled

Dr. Shibulal A
General Manager