



# KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/15176/2024

Dated: 18.05.2025

## NOTICE

**Sub:-** Tender No.T550(R)/ 2024 - Results of demonstration/ document evaluation & Date of pricebid opening – Iron Cot SS - reg.

**Ref:-** 1. Tender Inviting Notice published dated 27.08.2024  
2. Technical bid opening held on 28.09.2024 @ 11.00 A M  
3. Eligible list for demonstration published dated 27.03.2025  
4. Demonstration/ document evaluation of Iron Cot SS is conducted on 23.04.2025

I. The reports of technical document evaluation/ demonstration of Iron Cot SS as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models
1.	Iron Cot SS	M/s. STEEL INDUSTRIALS KERALA LIMITED	Custom/Steel Industrials

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s.Parkins Enterprises

Rejected model; MMM-61/Mercury medical makers

Reason for rejection:

Clause No	Technical Specification	Deviations noted
4	The top sheet shall be 20 gauge(1mm) thickness SS 304 sheet double bend of height minimum reinforced with channels of size minimum 75mm x 30mm flange welded inside the lengthwise frame bends.	49 x 25 mm. Not meeting the specification
5	The bed frame should be provided widthwise 3 nos of SS 304 angle stiffener of size minimum 35x35x2mm and 2 nos angle support size 50x35x2mm one at each head and leg.	Not meeting the specification
8	The leg frame at the head end should have a height of 1075mm and leg side should be at least 820mm.	1025 mm
9	Angular bracket of 3mm thickness shall be used for fixing the bed frame	Not meeting the specification

II. The online price bid of the eligible tenderer mentioned above will be opened on 21.05.2025 @ 11.AM at the e-tender portal.

**Dr. Vinay Goyal IAS**

Managing Director