



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

**EP Division**

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

**No: KMSCL/EP/10358/2025**

**Dated: 03.01.2026**

**NOTICE**

**Sub:-** Tender No.T579(R)/ 2025 - Cancellation Notice – Automated CPR System Model B- reg:-

**Ref:-** 1. Tender Inviting Notice published dated 29.07.2025

2. Technical bid opening held on 11.08.2025 @ 11.00 AM

3. Reports of demonstration/ document evaluation of Automated CPR System Model B recieved on 25.12.2025

I. The reports of technical document evaluation/ demonstration of various items as per reference no. 3rd cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers participated in the technical evaluation are as follows:

1. M/S Biomedical Engineering Company

Rejected model: Autopulse NXT/ Zoll Medical Corporation

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
12	Should be portable and weight and should be less than 5 Kg with a carrying case.	Heavier than the specified limit
14.a	Radiolucent Back Plate:01 set	Nil

II. In accordance with the Procurement norms ,the tender for KMSCL/EP/T579(R)/1394B/2025: Automated CPR System Model B stands cancelled

**Dr Arun P V**

General Manager