

**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/1608/2025

Dated: 21.10.2025

NOTICE

Sub:- Tender No.T565/ 2025 - Results of demonstration/ document evaluation & Date of pricebid opening – Diode Laser Triple Wave Length -reg:-

Ref:- 1. Tender Inviting Notice published dated 05.02.2025
2. Technical bid opening held on 01.03.2025 & 10.03.2025 @ 11.00 AM
3. Eligible list published on 03.06.2025
4. Reports of demonstation/ document evaluation of Diode Laser Triple Wave Length received on 25.09.2025

I. The reports of technical document evaluation/ demonstration of Diode Laser Triple Wave Length as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

| Sl. No | Equipment | Name of the Selected Bidder | Selected Models |
|--------|--------------------------------|------------------------------------|-----------------------------|
| 1. | Diode Laser Triple Wave Length | M/s. Kerala Surgical Equipment Co. | DXL Sakura Plus/Derma India |

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

A. M/s. Inscon Medical Soutlions PvtLtd

Rejected model: MBT- PACER ONE/Beijane Mega Beauty

Reasons for rejection:

| Clause No | Technical Specification | Deviations noted |
|-----------|--------------------------------|--|
| 4 | Repetition frequency – 1-20 Hz | 10Hz |
| 5 | Pulse width 10 -400 ms | 360 ms |
| | Remarks | The demonstrated model has variations in specifications with regard to repetition frequency and pulse width. |

II. The online price bid of the eligible tenderer mentioned above will be opened on ^{09.11.25} 05.11.2025 @ 11.AM at the e-tender portal.

Dr Arun P V

General Manager