



EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/7791/2025

Dated: 07.01.2026

**NOTICE**

**Sub:-** Tender No.T574R/ 2025 - Results of demonstration/ document evaluation & Date of pricebid opening – various equipment -reg;-

- Ref:-**
1. Tender Inviting Notice published dated 14.05.2025
  2. Technical bid opening held on 09.06.2025 @ 11.00 AM
  3. Eligible list published on 24.09.2025
  4. Demonstation/ document evaluation of Fume Hood is conducted on 18.11.2025
  5. Demonstation/ document evaluation of Vortex mixture is conducted on 27.11.2025
  6. Demonstation/ document evaluation of Sterilizer various types is conducted on 03.12.2025

I. The reports of technical document evaluation/ demonstration of Fume Hood as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models/OEM
1.	Fume Hood	M/s Labline Instruments	FC-3/LABLIN INSTRUMENTS
		M/s UNIVERSAL AGENCIES	YSI-225/YORK SCIENTIFIC
2.	Vortex Mixture	M/s.UNIVERSAL AGENCIES	MIX30PRO/MOXCARE
3.	Sterilizer Various Types	M/s UNIVERSAL AGENCIES	STERILIZER SMALL GIVA 1265 ,STERILIZER MEDIUM GIVA 1786 , STERILIZER LARGE GIVA 2086/GEETA INDUSTRIE

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

**1. KMSCL/EP/T574(R)/785A/2025: Fume Hood**

- A. M/s Yorco Sales Pvt Ltd  
 Rejected Model: YSL/YSI-225/ Yorco Sales  
 Reasons for rejection: Sample not produced

**2. KMSCL/EP/T574(R)/115/2025: Sterilizer Various Types**

- A. M/s. Mark Enterprises  
 Rejected Model: MSCD-03/Mark Enterprises  
 Reason for rejection: Sample not produced

B. M/s Zenotrex Medico Pvt. Ltd

Rejected Model: ZPL-A22R/ Zenotrex

Reason for rejection: Sample not produced

C. M/s Trivandrum Surgicals and Lab Supplies

Rejected Model: DUKE-240/BHARAT SURGICAL

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
	Should have industrial grade	
I.3	heaters with direct power connection	Not meeting the specification
II.3	Should have industrial grade heaters with direct power connection	Not meeting the specification
III.3	Should have industrial grade heaters with direct power connection	Not meeting the specification

III. The online price bid of the eligible tenderer mentioned above will be opened on 12.01.26 @ 11.AM at the e-tender portal.

**Dr Arun P V**  
General Manager