



## KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : [eptenders.kmscl@kerala.gov.in](mailto:eptenders.kmscl@kerala.gov.in)

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

No: KMSCL/EP/10401/2024

Dated: 18.07.2025

### NOTICE

**Sub:-** Tender No.T541/ 2024 - Cancellation Notice – Table Top Incubator (Compact Microbiological Incubator) - reg.

**Ref:-** 1. Tender Inviting Notice published dated 11.06.2024  
2. Technical bid opening held on 29.07.2024@ 11.00 A M  
3. Eligible list published on 01.01.2025  
4. Reports of demonstration/ document evaluation of Table Top Incubator (Compact Microbiological Incubator) received on 29.03.2025

I. The reports of technical document evaluation/ demonstration of Table Top Incubator (Compact Microbiological Incubator) as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

SI No	Equipment	Name of the Selected Bidder	Selected model
1.	Table Top Incubator (Compact Microbiological Incubator)	M/s Labline Instruments	BOD-1/LABLINE

1. M/s. Kerala surgical Equipment co .

Rejected model: PI-SPL D SS/AEON MEDITECH

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
2	Should have 18L capacity	Not meeting the specification
3	Should have LED display	Not meeting the specification
5	Dimensions: (D x W x H) Interior 11.4 x 7.1 x 12.2 in. (29 x 18 x 31 cm) approx. Dimensions: (L x W x H) 18.5 x 10.2 x 16.3 in. (47 x 26 x 41.5 cm) approx	Not meeting the specification
6	Temperature Range (Metric) +5°C to 50°C.	Not meeting the specification

2. M/s. UNIVERSAL AGENCIES

Rejected model:YSI-438D/YORK SCIENTIFIC

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
2	Should have 18L capacity	Not meeting the specification
3	Should have LED display	Not meeting the specification
5	Dimensions: (D x W x H) Interior 11.4 x 7.1 x 12.2 in. (29 x 18 x 31 cm) approx. Dimensions: (L x W x H) 18.5 x 10.2 x 16.3 in. (47 x 26 x 41.5 cm) approx.	Not meeting the specification
6	Temperature Range (Metric) +5°C to 50°C.	Not meeting the specification

II. In accordance with procurement norms the tender for **KMSCL/EP/T541/1425B/2024: Table Top Incubator (Compact Microbiological Incubator)** is cancelled

**Dr. Vinay Goyal IAS**  
Managing Director