



**KERALA MEDICAL SERVICES CORPORATION LTD**

(A Government Of Kerala Undertaking)  
 Thycaud.P.O., Thiruvananthapuram-14.  
 Tele Fax No. 0471-2945647  
 0471-2945646, 0471-2945600  
 E-mail id : [edrugs.kmscl@kerala.gov.in](mailto:edrugs.kmscl@kerala.gov.in)  
 CIN : U24233KL2007SGC021616  
 GSTIN: 32AADCK4029M1ZK

**Essential Drugs Division**

[www.kmscl.kerala.gov.in](http://www.kmscl.kerala.gov.in)

**No: KMSCL/ED/039/2025**

**Dated: 14.05.2026**

**Notice on Selected Bidders for Price Bid Opening**

**Sub:-** KMSCL – e-Tender for the Procurement of Test kits for NVBDCP 2024-25 & PLAN FUND 2025-26- -List of Selected/Rejected Bidders/ Products for Price Bid Evaluation – Published – Reg.

**Ref:-** 1. E-Tender No- KMSCL/ED/TEST KITS-PLAN& NVBDCP/RC/2026 /001 dtd 31.01.2026.  
 2. This office Notice No-KMSCL/ED/39/2025 dtd.05.03.2026.

The Technical bids submitted in response to the e-tender referred above for the Procurement of **Test kits for NVBDCP 2024-25 & PLAN FUND 2025-26** along with the clarifications/complaints received against the notice 2<sup>nd</sup> cited were scrutinized.

The list of selected/rejected bidders/products based on the scrutiny of the documents/complaints and qualitative evaluation of samples is published below.

The price bids of all the qualified bidders will be opened and evaluated from **11.30am** onwards on **15.05.2026**.

Sl No	Drug Code	Item Name	Strength	Unit	Decisions taken
<b>1. M/s. Meril Diagnostics Private Limited, Gujarat</b>					
1	D26044	DENGUE RAPID KITS - COMBO (IGG+IGM+NS1)	1Test	1 No	The samples of the following items submitted for qualitative evaluation as per clause 6.16 of the tender document is <b>rejected</b> by the technical committee for the reasons mentioned;  1. D26044 -DENGUE RAPID KITS – COMBO (IGG+IGM+NS1) ( <b>Sensitivity is poor</b> ).  2. DENGUE NS1 ANTIGEN ELISA KIT ( <b>Sensitivity is poor</b> ).  <b>The offer of the other item is accepted.</b>
2	D26010	RDT MALARIA BIVALENT KITS	1Test	1 No	
3	D26038	DENGUE NS1 ANTIGEN ELISA KIT	1 Test	1 No	
<b>2. M/s. S D Biosensor Healthcare Pvt.Ltd, Haryana</b>					
1	D26044	DENGUE RAPID KITS - COMBO (IGG+IGM+NS1)	1Test	1 No	The samples of the following items submitted for qualitative evaluation as per clause 6.16 of the tender document is <b>rejected</b> by the technical committee for the reasons mentioned;  1. D26044 -DENGUE RAPID KITS – COMBO (IGG+IGM+NS1) ( <b>Sensitivity is poor</b> ).  2. D26038- DENGUE NS1 ANTIGEN ELISA KIT ( <b>Sensitivity is poor</b> ).  <b>The offer of the other item is accepted.</b>
2	D26010	RDT MALARIA BIVALENT KITS	1Test	1 No	
3	D26038	DENGUE NS1 ANTIGEN ELISA KIT	1 Test	1 No	

3. M/s.Oscar Medicare Private Limited, New Delhi					
1	D26044	DENGUE RAPID KITS - COMBO (IGG+IGM+NS1)	1Test	1 No	The samples of the following items submitted for qualitative evaluation as per clause 6.16 of the tender document is <b>rejected</b> by the technical committee for the reasons mentioned.  1.D26044 -DENGUE RAPID KITS – COMBO (IGG+IGM+NS1) ( <b>Sensitivity is poor</b> ). 2.D26038- DENGUE NS1 ANTIGEN ELISA KIT ( <b>Sensitivity is poor</b> ).  <b>The offer of the other item is accepted.</b>
2	D26010	RDT MALARIA BIVALENT KITS	1Test	1 No	
3	D26038	DENGUE NS1 ANTIGEN ELISA KIT	1 Test	1No	
4. M/s. Med source Ozone Biomedicals Private Limited, Haryana					
1	D26044	DENGUE RAPID KITS - COMBO (IGG+IGM+NS1)	1Test	1No	The samples of the following items submitted for qualitative evaluation as per clause 6.16 of the tender document is <b>rejected</b> by the technical committee for the reasons mentioned;  D26044 -DENGUE RAPID KITS – COMBO (IGG+IGM+NS1) ( <b>Sensitivity is poor</b> ).  <b>The offer of the other items is accepted.</b>
2	D26010	RDT MALARIA BIVALENT KITS	1Test	1 No	
3	D26038	DENGUE NS1 ANTIGEN ELISA KIT	1 Test	1 No	

Sd/-

**Dr Arun P V**  
General Manager