



KERALA MEDICAL SERVICES CORPORATION LTD

(A Government Of Kerala Undertaking)

Thycaud.P.O., Thiruvananthapuram-14.

Tele Fax No. 0471-2945647

0471-2945646, 0471-2945600

E-mail id : eptenders.kmscl@kerala.gov.in

CIN : U24233KL2007SGC021616

GSTIN: 32AADCK4029M1ZK

EP Division

www.kmscl.kerala.gov.in

No: KMSCL/EP/4414/2025

Dated: 13.11.2025

NOTICE

Sub:- Tender No.T569/ 2025 - Results of demonstration/ document evaluation & Date of pricebid opening – various equipment- reg.

- Ref:-**
1. Tender Inviting Notice published dated 19.03.2025
 2. Technical bid opening held on 07.04.2024 @ 11.00 A M
 3. Eligible list for demonstration published dated 27.08.2025
 4. Demonstration/ document evaluation of various equipment is conducted on 03.11.2025

I. The reports of technical document evaluation/ demonstration of various equipment as per reference no. 4th cited above were submitted by the technical committee appointed for the purpose. The details of the tenderers qualified in the technical evaluation are tabulated as follows:

Sl. No	Equipment	Name of the Selected Bidder	Selected Models/ OEM
1.	IV Stand (Drip Stand)	M/s. Kerala Surgical Equipment Co.	MED-50/Medik
		M/s. Parkins Enterprises	MM 003/MERCURY
2.	X-Ray Film Lobby - LED- Single Film	M/s. UNIVERSAL AGENCIES	SPARX XVB 14/HARDIK
		M/s. Kerala Surgical Equipment Co.	Lovanni LED -1/ Lovaani Impex
		M/s. Parkins Enterprises	LED 15W/:BIO-X
3.	X-Ray Film Lobby - LED - Double Film	M/s. UNIVERSAL AGENCIES	SPARX XVB 14/HARDIK

II. The other tenderers participated in the tender are rejected due to the reasons mentioned against their name below;

1. KMSCL/EP/T569/365/2025: IV Stand (Drip Stand)

A. M/s. U K SURGICALS

Rejected Model: MW-32/BHNUKS101/ UK SURGICALS

Reason for rejection: Sample not produced.

B. M/s. BHN Surgicals Private Limited

Rejected model: MW-32/BHN

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
3	Pre treated and Powder coated.	Not meeting the specification

2. KMSCL/EP/T569/412A/2025: X-Ray Film Lobby - LED- Single Film

A. M/s. X ray medico surgical engineering ancillary

Rejected model: XMSEA/X-ray Medico

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
4	Should have minimum 10000 Lux output adjustable.	Not meeting the specification

3. KMSCL/EP/T569/412B1/2025: X-Ray Film Lobby - LED - Double Film

A. M/s. Kerala Surgical Equipment Co

Rejected model: Lovanni LED -2/ Lovaani Impex

Reason for rejection: Sample not produced

B. M/s. Parkins Enterprises

Rejected model; LED 30W: BIO-X

Reason for rejection:

Clause No	Technical Specification	Deviations Noted
9	Individual light controls for each plate.	Not meeting the specification

III. The online price bid of the eligible tenderer mentioned above will be opened on 18.11.2025 @ 11.AM at the e-tender portal.

Dr Arun P V
General Manager